

## **CORRIGENDUM to the CONTRACT NOTICE and TENDER DOSSIER**

---

**Contract Title:** Provision for Approved Laboratory Test Result

**Publication Reference:** SIHHAT/2021/SUP/INT/01

---

DOC: Document

ART: Article

ITT: c4b\_itt\_en [Instructions to Tenderers]\*

DC: Draft contract

TS: c4f\_annexiitechspeciitechoffer\_en [Annex II + III: Technical Specifications + Technical Offer]\*

*\* In-parenthesis parts show the title inside the documents.*

---

Further to the requests received from the tenderers, the following clarifications are provided.

#	DOC	ART/IT EM	FORMER TEXT / DOCUMENT(S)	SHALL READ AS NEW TEXT / DOCUMENT(S)
<b>Instruction of Tenderes (ITT)</b>				
1	ITT	1/ 1.1.	....Collecting samples from patients shall be performed on every working day until 11.30. The samples shall be ready for testing/examination at the laboratory between 11.30 a.m. and 14.00 and results shall be reported by 14.00 on the next day (within maximum 24 hours) (excluding thalassemia and microbiological culture/antibiogram tests)....	....Samples shall be received from patients until 11.30 a.m. on every working day. The Contractor shall collect the samples, which are received at the laboratory, between 11.30 a.m. and 14.00 p.m. and report the results (except for thalassemia and microbiological culture/antibiogram tests) by 14.00 p.m. on the next day....
<b>Draft Contract (DC)</b>				
2	DC	1 / 1.1.	....Collecting samples from patients shall be performed on every working day until 11.30. The samples shall be ready for testing/examination at the laboratory between 11.30 a.m. and 14.00 and results shall be reported by 14.00 on the next day (within maximum 24 hours) (excluding thalassemia and microbiological culture/antibiogram tests)....	....Samples shall be received from patients until 11.30 a.m. on every working day. The Contractor shall collect the samples, which are received at the laboratory, between 11.30 a.m. and 14.00 p.m. and report the results (except for thalassemia and microbiological culture/antibiogram tests) by 14.00 p.m. on the next day....
<b>Annex II + III: Technical Specifications + Technical Offer</b>				
3	TS	4 / 4.7.	Collecting samples from patients shall be performed on every working day until 11.30. The samples shall be ready for testing/examination at the laboratory between 11.30 a.m. and 14.00 and results shall be reported by 14.00 on the next day (within maximum 24 hours) (excluding thalassemia and microbiological culture/antibiogram tests).	Samples shall be received from patients until 11.30 a.m. on every working day. The Contractor shall collect the samples, which are received at the laboratory, between 11.30 a.m. and 14.00 p.m. and report the results (except for thalassemia and microbiological culture/antibiogram tests) by 14.00 p.m. on the next day
4	TS	7./7.12.	The devices to be installed should not complete 10 (ten) years by the end of the term of the contract. The Contractor shall provide catalogues for technical information about the devices to be installed and kits to be worked with, clearly indicate names of manufacturing companies and this information shall be presented in the tender dossier.	The devices to be installed should not complete 13 (thirteen) years by the end of the term of the contract. The Contractor shall provide catalogues for technical information about the devices to be installed and kits to be worked with, clearly indicate names of manufacturing companies and this information shall be presented in the tender dossier.

5	TS	8./8.2.	The Contractor should make regular internal quality controls for all tests on at least 2 (two) levels every day and be a member to an external quality control program (of approved organizations) for all tests within the scope of work.	The Contractor should make regular internal quality controls for all tests (excluding Microscopic tests and microbiological culture/antibiogram tests) on at least 2 (two) levels every day and be a member to an external quality control program (of approved organizations) for all tests within the scope of work.
6	TS	8./8.5	Sample acceptance and rejection criteria should be defined and implemented. The purpose is to inform both the laboratory personnel and other related E/MHC personnel about the rules on sample acceptance to obtain accurate results from the tests and minimize preanalytical errors in particular. A Test Manual including times of sample processing, sample types, information about the tests which require preliminary preparation, sample acceptance or rejection criteria etc. shall be available. The Contractor shall be responsible for translating and distributing introductory brochures prepared by the Administration.	Sample acceptance and rejection criteria should be defined and implemented. The purpose is to inform both the laboratory personnel and other related E/MHC personnel about the rules on sample acceptance to obtain accurate results from the tests and minimize preanalytical errors in particular. A Test Manual including times of sample processing, sample types, information about the tests which require preliminary preparation, sample acceptance or rejection criteria etc. shall be available. The Contractor shall be responsible for translating and distributing related documents. Quantities of these documents will be provided by the contractor due to the requirements. All brochures/instructions and training documents should be in English, Arabic and Turkish.
7	TS	11./11.7	In cases tests cannot be performed at a laboratory for a temporary period due to technical reasons, the Contractor should make an agreement with another laboratory/laboratories (other than the previously contracted subcontractor(s)) once approved by the Inspection and Acceptance Commission provided that test conclusion times and service flow are not disrupted. If the tests to be sent to an external laboratory are among those which cannot be run by the company at our laboratory on routine basis, conclusion times of the contracted external laboratory shall be applied. The laboratory/laboratories to be contracted by the Contractor should work with the current parameters using the same methods, run regular internal and external quality controls and be a member of an external quality control program for all current tests within the scope of work. It should also have a Biochemistry and Microbiology Laboratory License issued by the Ministry of Health. The Contractor is required to submit documents and agreements which demonstrate the required qualifications for this laboratory for approval of the Inspection and Acceptance Commission before starting the work. All tests to be run should be transferred to the LOS.	In cases tests cannot be performed at a laboratory for a temporary period due to technical reasons, the Contractor can work with another laboratory/laboratories once approved by the Inspection and Acceptance Commission provided that test conclusion times and service flow are not disrupted. If the tests to be sent to an external laboratory are among those which cannot be run by the company at our laboratory on routine basis, conclusion times of the contracted external laboratory shall be applied. The laboratory/laboratories to be contracted by the Contractor should work with the current parameters using the same methods, run regular internal and external quality controls and be a member of an external quality control program for all current tests within the scope of work. It should also have a Biochemistry and Microbiology Laboratory License issued by the Ministry of Health. The Contractor is required to submit documents and agreements which demonstrate the required qualifications for this laboratory for approval of the Inspection and Acceptance Commission before starting the work. All tests to be run should be transferred to the LOS.

8	TS	12.3/ 12.3.1.1.	The device should be able to operate with complete blood sample out of the primary tube without requiring any preliminary processing or taking the cap off.	The device should be able to operate with complete blood sample out of the primary tube without requiring any preliminary processing.
---	----	--------------------	---	---

**All other terms and conditions of the Contract Notice and Tender Dossier remain unchanged. The above alterations and /or corrections are integral part of the Contract Notice and Tender Dossier.**