

CLARIFICATION No: 1

to the TENDER DOSSIER

Publication Ref: SIHHAT/2018/SUP/INT/04

Subject: Supply of Furniture, Medical Equipment, Devices and Consumables for Migrant Health Centres

Location –Europe (non EU/Turkey)

The following clarification is made to the tender dossier.

CONTRACT NOTICE	
Question 1:	<p><u>Article 16.3:</u> If a tenderer submit financial offer more than one Lot, so the average annual turnover will be determined over the highest bid price for offered lots? Or the average annual turnover will be calculated over the total amounts of lots offered more than one lot?</p>
Answer 1:	<p>For tenderers submitting tenders for more than one lot, the average annual turnover of the tenderer in the last three financial years must exceed the cumulative amount of the financial offers of all the lots for which the tenderer submitted tenders.</p>
Question 2:	<p><u>Article 16.3:</u> In contract notice Item 16.3.b it says "The tenderer has delivered supplies under at most two contracts with a budget of at least one-fourth ($\frac{1}{4}$) of the financial proposal". We want to clarify the term "contracts". Can we consider invoices as a contract?</p>
Answer 2:	<p>Contract is a legal document that states and explains a formal agreement between two different people or groups, or the agreement itself. According to article 16.3.b only contract(s) is acceptable. Invoices are supporting/proof documents for contracts and tenderer has to give invoice(s) with contract(s).</p>
Question 3:	<p><u>Article 16.3:</u> If a tenderer submit financial offer more than one Lot, so the tenderer has delivered supplies under at most 2 contracts will be determined over the highest bid price for offered lots? Or the tenderer has delivered supplies under at most 2 contracts will be calculated over the total amounts of lots offered more than one lot?</p>
Answer 3:	<p>Tenderers submitting tenders for more than one lot must have delivered similar supplies either (i) under a single contract with budget of at least the cumulative amount of the financial offers of all the lots which the tenderer submitted tenders, or (ii) under at most 2 contracts with budgets of at least the amounts of the respective financial offers of the lots for which the tenderer submitted tenders.</p>

Question 4:	<i>Article 15. Period of implementation of tasks</i> Delivery time is too short. Is it possible to be a 90-120 days?
Answer 4:	Please follow Article 15 “Period of implementation of tasks” of Contract Notice and Article 13 “Programme of implementation of tasks” of Special Conditions.
SPECIAL CONDITIONS	
Question 5:	Could you give us some photos or plottings for lot 1 and lot 2?
Answer 5:	Please follow Item 7.1 under Article 7 “Supply of documents” and Item 14.1 under Article 14 “Contractor’s drawings” of Contract Notice and Please see the Corrigendum No: 2 to Tender Dossier.
Question 6:	Which documents are required to participate in the tender?
Answer 6:	Please follow Article 18 “Language of the procedure” of Contract Notice.
Question 7:	Is there a Turkish translation of the tender documents? If so, how can we provide it?
Answer 7:	Please follow Article 21 “Language of the procedure” of Contract Notice and Article 2.1 “Language of the Contract” of Special Conditions.
Question 8:	According to Item 32.7 Special Conditions “The warranty shall remain valid for 2 (two) years after provisional acceptance.” However Item 9.1.16.1 under Technical Specification Lot 9 “The devices (system) will be guaranteed at least 5 years...” which one is correct?
Answer 8:	Please see the Corrigendum No: 2 to Tender Dossier.
ANNEX II+III: TECHNICAL SPECIFICATIONS + TECHNICAL OFFER	
Question 9:	General Requirements Item 3.6.1, registration is required to National Information Bank of Medicine and Medical Device. However, Furniture, mentioned at LOT-1 is not related with Medicine or Medical Device. So they can not be registered at that platform. Please exclude LOT-1 from item 3.6.1.
Answer 9:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 10:	General Requirements Item 3.6.2, The tenderer companies shall give documents certifying company identification number indicating that they are registered with the TITUBB if they are manufacturers and / or importers of the products offered in this scope; the dealer identification numbers if they are dealers together with the offer or it shall be asked before contract signature. Are the lot 1 and lot 2 in the scope of T.C. Ministry of Health TITUBB approval?
Answer 10:	Please see the Corrigendum No: 2 to Tender Dossier.

Question 11:	General Requirements Item 3.6.3, it is asking for hardness test values for surgical instruments. We want to clarify that, is a declaration enough from manufacturer that stated all products comply with hardness requirement? If not which documents we need to provide?
Answer 11:	The declarations is enough when tenderer submit their tender dossier. However according to Article 29 Annex I: General Conditions and Special Conditions, the supply shall include all necessary documents as specified herein such as operating and maintenance manuals, drawings, material certificates, conformity certificates, test certificates, certificates of origin, planning, packing lists and others as necessary before Provisional Acceptance.
Question 12:	General Requirements Item 3.6.8. The representative company shall document its technical service facilities and infrastructure. The company shall submit the Service Location Qualification Certificate, which it has received from TSE (according to TSE 13011 and TSE 12426 standards), or an international standart with its offer. Is the TSE certificate required to lot 1 and lot 2?
Answer 12:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 13:	General Requirements Item 3.6.8. The representative company shall document its technical service facilities and infrastructure. The company shall submit the Service Location Qualification Certificate, which it has received from TSE (according to TSE 13011 and TSE 12426 standards), or an international standart with its offer. Are the TSE 13011-TSE 12526 certificates required to lot 1 and lot 2? Or they're in the scope of medical devices?
Answer 13:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 14:	General Requirements Item 3.6.8, The documents required in Article 3.6.8 is for medical products. It does not apply to furniture. Service location qualification certificate is legally determined as TSE 12487. We ask that removal of TSE 13011 and TSE 12426 documents for furniture from specification and change as TSE 12487 or equivalent.
Answer 14:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 15:	General Requirements Item 3.6.8, The documents required in Article 3.6.8 is for medical products. It does not apply to furniture. Service location qualification certificate is legally determined as TSE 12487. We ask that removal of TSE 13011 and TSE 12426 documents for furniture from specification and change as TSE 12487 or equivalent. Suggested clause "The representative company shall document its technical service facilities and infrastructure. The company shall submit the Service Location Qualification Certificate, which it has received from TSE (according to TSE 13011 and/or TSE 12426 standards), or an international standart with its offer"

Answer 15:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 16:	Item 1.1 to 1.19, Technical specification is not enough for Furniture, LOT-1. In order to offer correct solution, please provide indicative technical drawings or photos for all items at LOT-1, except 1.17 and 1.19. Indicative /reference drawing or photo is usually supplied at Furniture tenders by contractor authority. So, one of them is very helpful to collect equivalent solution from the participants.
Answer 16:	Please follow Article 1.1 to 1.19 and subtitles of Technical Specifications. Please see the Corrigendum No: 2 to Tender Dossier.
Question 17:	Item 1.1.4, “Artificial leather or fabric should be used in the coating,” Suggested clause “It shall be manufactured with Fabric: 100% Polyester, Artificial Leather: 18% Polyester- 80% PVC - 2% Polyurethane min: 400 gr / m2 max: 600 gr / m2 and it shall be possible to be clean it with a damp cloth.”
Answer 17:	Please follow Article 1.1.4 of Technical Specifications.
Question 18:	Item 1.1.5, “It must be a sponge cut on metal frame and metal profile,” Suggested clause “it shall be metal skeleton on cast sponge”
Answer 18:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 19:	Item 1.1.6, “Seat height will be 78 ± 2 cm.” Suggested clause “The seat height shall be $90 \text{ cm.} \pm 3 \text{ cm.}$ ”
Answer 19:	Please follow Article 1.1.6 of Technical Specifications.
Question 20:	Item 1.1.7, “Seat depth will be 70 ± 2 cm.” Suggested clause “The seat depth shall be $62 \text{ cm.} \pm 3 \text{ cm.}$ ”
Answer 20:	Please follow Article 1.1.7 of Technical Specifications.
Question 21:	Item 1.1.8, “Seat width will be 155 ± 2 cm.” Suggested clause “The seat depth shall be $160 \text{ cm.} \pm 5 \text{ cm.}$ ”
Answer 21:	Please follow Article 1.1.8 of Technical Specifications.
Question 22:	Item 1.1.9, “The seating depth will be 50 ± 2 cm.” Suggested clause “The seat depth shall be $62 \text{ cm.} \pm 3 \text{ cm.}$ ”
Answer 22:	Please follow Article 1.1.9 of Technical Specifications.
Question 23:	Item 1.1.10, “The seating surface width will be 60 ± 2 cm.” Suggested clause “The seating surface width shall be $104 \text{ cm.} \pm 5 \text{ cm.}$ ”
Answer 23:	Please follow Article 1.1.10 of Technical Specifications.

Question 24:	Item 1.1.11, "It shall be 38 cm above the sitting surface of the upper backrest." Suggested clause "It shall be 45cm ± 3 cm. above the sitting surface of the upper backrest."
Answer 24:	Please follow Article 1.1.11 of Technical Specifications.
Question 25:	Item 1.1.12, Made of 40 x 40 / 1.5 mm profiler. Hard plastic material should be used on the surfaces of the feet that come into contact with the ground." Suggested clause "Footwork; 21x46x2 mm and intermediate weft Ø60x2 mm steel pipes."
Answer 25:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 26:	Item 1.1.13, "In accordance with body ergonomic requirements, the session surface will be used 8 cm 32 dns ridge surface 5 cm 32 dns." Suggested clause "In accordance with body ergonomic requirements, the session surface will be used 8 cm 50± 5 dns (kg/m3) ridge surface 5 cm 50± 5 dns (kg/m3)"
Answer 26:	Please follow Article 1.1.13 of Technical Specifications.
Question 27:	Item 1.1.14, "Upholstery fabric and artificial polyurethane leather will be used on backrest and sitting surface." Specifications 1.2.4 and 1.10.4 as well as upholstery fabric or artificial polyurethane leather. Artificial leather and fabric expression was used in article 1.1.14. The coating is not both artificial leather and fabric.
Answer 27:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 28:	Item 1.1.15, "Skeleton floor-surface 20 x 20 x 1,5 mm profile back-back surface 20 x 30 x 1,5 mm profile will be made." Suggested clause "Carrier traverse shall be made of Ø60 * 2 pipes and 3 pieces of 3 mm connection plate shall be welded to the end of the traversing oval pipe legs of 46-21 x 2 mm by special welding method. The legs welded with the carrier traverse shall be covered with chrome coating. Parts made of PA (polyamide) material shall be used."
Answer 28:	Please follow Article 1.1.15 of Technical Specifications.
Question 29:	Item 1.2, Do you require height adjustment at item 1.2, Working Chair? Also, what is the meaning of "Seat Belt" at 1.2.11?
Answer 29:	Please follow Article 1.2 and subtitles of Technical Specifications. Please see the Corrigendum No: 2 to Tender Dossier.
Question 30:	Item 1.2.4, "Artificial leather or fabric should be used. A 32 kg / m ³ gray cut sponge should be used. Plastic armrests and star stand should be manufactured. Damortion, mechanism and plastic wheel should be used." Suggested clause "It shall be fabric: 100% Polyester, Artificial Leather: 18% Polyester - 80% Pvc - 2% Polyurethane min: 400 gr / m ² max: 600 gr

	/ m ² it shall be possible to be cleaned with a damp cloth and metal skeleton On Cast sponge.”
Answer 30:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 31:	Item 1.2.4, Is it possible that Casting sponge or cut sponge?
Answer 31:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 32:	Item 1.2.5, “The seating depth will be 46 ± 2 cm.” Suggested clause “The seating depth shall be 48 ± 3 cm.”
Answer 32:	Please follow Article 1.2.5 of Technical Specifications.
Question 33:	Item 1.2.6, “The seating surface width will be 48 ± 2 cm.” Suggested clause “The seating depth shall be 52 ± 3 cm.”
Answer 33:	Please follow Article 1.2.6 of Technical Specifications.
Question 34:	Item 1.2.7, “The height from the seating surface of the upper backrest will be 48 ± 2 cm.” Suggested clause “The height from the seating surface of the upper backrest will be 46 ± 2 cm.”
Answer 34:	Please follow Article 1.2.7 of Technical Specifications.
Question 35:	Item 1.2.8, “The backrest width will be 45 cm widest.” Suggested clause “The backrest width will be 48 ± 3 cm. widest.”
Answer 35:	Please follow Article 1.2.8 of Technical Specifications.
Question 36:	Item 1.2.10, “The sponge will be a cut sponge with a density of at least 32 kg / m ³ .” Suggested clause “The seat made of monobloc body shall be made of special aluminum molds with density of 50 ± 5 (density, kg / m ³) according to body ergonomics on steel carcass.”
Answer 36:	Please follow Article 1.2.10 of Technical Specifications.
Question 37:	In items 1.2.11 and 1.10.11, a term "seat belt" used to describe an article of furniture. Do you mean the construction material/skeleton of chairs? If not could you provide detailed description about "seat belt"?
Answer 37:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 38:	In items 1.2.11 and 1.10.11, “safety belts” were not understood. We request clarification on the statement.
Answer 38:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 39:	Item 1.2.12, “It should have a star-like appearance and subjected to a black impact and load test. The main frame should be reinforced with press. The

	<p>seat wheel is made of plastic and has been subjected to a crash and overtravel test.”</p> <p>Black impact should be break impact. This expression is used in the automotive sector. It is not a valid method for this product.</p> <p>Load test: The load rating at which the wheels and the plastic assembly can be lifted is given. The test report can not be requested in this type of products.</p> <p>Crash and overtravel: the desired characteristics do not include the necessity of making human health and proof values.</p> <p>For these reasons, we request that this clause be removed.</p>
Answer 39:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 40:	<p>Item 1.2.12, “The foot on the seat is made of five-spoke and five-wheel plastic. It should have a star-like appearance and subjected to a black impact and load test. The main frame should be reinforced with press. The seat wheel is made of plastic and has been subjected to a crash and overtravel test”</p> <p>Suggested clause “Wheels shall be produced in accordance with soft floors from polyamide material. Armrests shall be made of polypropylene (PP) material. Plastic Starch and PA6 (Polyamide) wheels made from materials that meet the PAH criteria shall provide contact with the wheel seat.”</p>
Answer 40:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 41:	<p>Item 1.2.13, “At least 2.2 mm thick should be molded. The swing must have a moving twin-spoke mechanism.”</p> <p>Suggested clause “It shall have an ergonomic structure, the Monoblock injection cast sponge with a thickness of 6 cm shall be at a density of 50 (density, kg / m³)”</p>
Answer 41:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 42:	Item 1.3, We need to some photo or plotting of item 1.3. Desk bench.
Answer 42:	Please follow Article 1.3 and subtitles of Technical Specifications. Please see the Corrigendum No: 2 to Tender Dossier.
Question 43:	<p>Item 1.3, The height related to the product is given as 110 cm. Is a table with a height separator used in this product? Are there any visual of this product?</p> <p>Do you have the possibility to provide sample visuals for the products used throughout the tender dossier?</p>
Answer 43:	Please follow Article 1.3 and subtitles of Technical Specifications. Please see the Corrigendum No: 2 to Tender Dossier.
Question 44:	Item 1.3.5, there are 2 side tables and front table. Is the desk bench u-shaped or l-shaped? We need to some photo or plotting of desk bench.

Answer 44:	Please follow Article 1.3 and subtitles of Technical Specifications. Please see the Corrigendum No: 2 to Tender Dossier.
Question 45:	Item 1.3.5, it describes a dimension "40 and 40 x 1.5 x 40 for the profile used in the foot", on that sentence 40 used 2 times. Is it an error?
Answer 45:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 46:	Item 1.3.5, "Table top table 25 mm top plate melamine 2 side table and front table 25 mm chipboard melamine material on the plate will be manufactured. 40 and 40 x 1.5 x 40 for the profile used in the foot and metal corner components will be used for brackets." Suggested clause "The table top and the top table shall be covered with a 0.6 mm lartinate of double sided 30 mm chipboard. The feet shall be 30mm melamine coated chipboard. The front panel shall be 18mm melarnine coated chipboard."
Answer 46:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 47:	Item 1.3.6, "The decorative parts used in front of the table will be melamine on 25 mm chipboard." Suggested clause "The four sides of the top table, the long side of the table top 1, the short sides of the feet 1, the long side and the front side 4 of the front panels shall be covered with 2 mm pvc. The counter shall be installed with steel fixtures. Two plastic cable lugs of diameter Ø60 mm shall be used to provide cable passage. It shall be capable of being raised from the ground with plastic lugs."
Answer 47:	Please follow Article 1.3.6 of Technical Specifications.
Question 48:	Item 1.4.4, "It should be at 80 x 40 x 160 cm" Suggested clause "It should be at $80 \pm 3 * 40 \pm 2 * 155 \pm 5$ cm."
Answer 48:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 49:	Item 1.4.5, "The library top plate will be 30 mm, side plate, partition, shelf, bottom plate and covers 18 mm, backing 8 mm double face melamine resin impregnated decor paper coated chipboard. with ± 2 mm tolerance" Suggested clause "The top surface shall be melamine coated on a 30 mm chipboard. The substrate should consist of a bottom plate, two side tables and a fixed shelf, joined by saddle stitchers and mounted on a 8 mm chipboard with a melamine coated backing grooved system. Lower, side and shelves 18 mm the bottom covers shall be melamine coated 18 mm chip board."
Answer 49:	Please follow Article 1.4.5 of Technical Specifications.
Question 50:	Item 1.4.7, "The library top, bottom and library covers are 4 edges 1 mm, all other edges visible after mounting are 1 mm PVC with ± 2 mm tolerance" Suggested clause "The front edge of the 30 mm top table shall be 2 mm long and the other edges shall be 0.440 mm PVC tape. The bottom edge of the

	bottom table, the long edge of the side table 2, the front edge of the fixed shelf, the 4 sides of the movable shelves shall be PVC tape. 4 sides shall be 2 mm PVC tape.”
Answer 50:	Please follow Article 1.4.7 of Technical Specifications.
Question 51:	Item 1.4.8, “The general construction will be made with MiniFix connectors. The handle will be 128 mm in size and (spring) chrome plated. The movable shelves will be mounted on four shelf pins. The cabinet leg should be metal provided and fixed. The doors should be locked with one lock.” Suggested clause “The cupboard shall be mounted with steel minifiks. The cup shall be raised with 4 equilibrium aluminum legs. There shall be 2 flat hinges on the doors. Each door shall have a handle with 128mm between two holes. The movable shelves shall be mounted on the metal shelf pins attached to the side tables.”
Answer 51:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 52:	Item 1.4.8, 1.6.7 and 1.9.7, The handle will be 128 mm in size and (spring) chrome plated. Precise measurement is used as in the general specification, and no tolerance values are given. Tolerance values are not specified for generic specimens. We demand that the dimension of 128 mm should be changed to at least 110 mm.
Answer 52:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 53:	Item 1.5.4, “seat and back should be plastic material” Polyurethane material is preferred in new generation productions in terms of strength and flexibility. We ask that the relevant article be corrected as follows. “seat and back should be plastic or polyurethane material”
Answer 53:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 54:	Item 1.5.4, “Four-legged metal frame, seat and back should be plastic material. Seating surface and back shall be cast sponge, should be plastic shoe. Artificial leather or fabric covered.” Suggested clause “Footwork; 30 * 15 * 1.5 mm oval pipe and interlining Ø16 * 1,5 mm pipe, Ergonomic structure, 2.5 cm thick seat and backrest injection casting sponge shall be 50 (density, kg/m ³ density Fabric: 100% Polyester, Artificial Leather: 18% Polyester - 80% Pvc - 2% Polyurethane Min: 400 gr / m ² max: 600 gr / m ² • It shall be cleaned with a damp cloth. Seat and back shall be cast sponge.”
Answer 54:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 55:	Item 1.5.6, “The depth of seating will be 42 ± 2 cm.”

	Suggested clause “The depth of seating shall be 59 ± 4 cm.”
Answer 55:	Please follow Article 1.5.6 of Technical Specifications.
Question 56:	Item 1.5.7, “The seating surface width will be 46 ± 2 cm.” Suggested clause “The seating surface width shall be 48 ± 3 cm.”
Answer 56:	Please follow Article 1.5.7 of Technical Specifications.
Question 57:	Item 1.5.8, “The height of the top of the backrest from the seating surface shall be 35 cm.” Suggested clause “The height of the top of the backrest from the seating surface shall be 57 ± 3 cm.”
Answer 57:	Please follow Article 1.5.8 of Technical Specifications.
Question 58:	Item 1.5.9, “The back width will be 46 ± 2 cm.” Suggested clause “The back width shall be 48 ± 2 cm.”
Answer 58:	Please follow Article 1.5.9 of Technical Specifications.
Question 59:	Item 1.5.10, “At least $45 \text{ kg} / \text{m}^3$ density foam sponge will be.” Suggested clause “It shall be ergonomic, it shall be 2,5 cm thick for the seat and backrest and shall be made of injection casting sponge of 50 dns (kg/m^3).
Answer 59:	Please follow Article 1.5.10 of Technical Specifications.
Question 60:	Item 1.5.11, “Feet and wheels will be $15 \times 30 \times 1,5 \pm 2$ mm oval pipe” Suggested clause “Footwork; $30 * 15 * 1.5$ mm oval pipe, and interlining $016 * 1.5$ mm pipe.”
Answer 60:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 61:	Item 1.5.11, There isn’t any product which is convenient to technical specifications. Word of wheel could be written irrelevently?
Answer 61:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 62:	Item 1.5.11, In the market, Education Chairs are supplied without wheels. According to technical specification 1.5.11, “feet and wheels” are described. Please clarify that wheels are necessary or not for item 1.5, Education Chairs. Also, do you need blackboard which is attached to Education Chair (item 1.5)?
Answer 62:	Please follow Article 1.5 and subtitles of Technical Specifications. Please see the Corrigendum No: 2 to Tender Dossier.
Question 63:	Item 1.5.12, “Skeletal Seat, The product consists of two parts as back, will be injection printing hard Plastic material.” Suggested clause “Seat plastic and backing plastics shall be manufactured in special molds made of flexible polypropylene (PP) according to PAH

	criteria. Arinrests shall be made of polypropylene (PP) material in accordance with PAH criteria.”
Answer 63:	Please follow Article 1.5.12 of Technical Specifications.
Question 64:	Item 1.5.13, “Artificial leather 2% polyurethane, 20% Polyester, 80% pvc, 450 gr / m2 will be. Fabric Polyester or polyolefin, at least 200 gr / m2, at least 20.000 will be resistant to abrasion.” Suggested clause “It shall be fabric: 100% Polyester, Artificial Leather: 18% Polyester - 80% Pvc - 2% Polyurethane min: 400 gr / m2 max: 600 gr / m2 and it shall be able to be cleaned with a damp cloth.”
Answer 64:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 65:	Item 1.5.14, “Profile pipe will be 15 x 30 x 1,5 mm oval pipe (feet).” Suggested clause “The profile pipe for feet shall be of 15 x 30 x 1.5 mm oval pipe.”
Answer 65:	Please follow Article 1.5.14 of Technical Specifications.
Question 66:	Item 1.5.16, “Metals shall be at least 40 micron electro-static painted.” Suggested clause “It shall be stoved with 40 micron electrostatic powder paint and fired at 200°C.”
Answer 66:	Please follow Article 1.5.16 of Technical Specifications.
Question 67:	Item 1.6, the dressing closet is for how many people?
Answer 67:	For 2 persons. Please see the Corrigendum No: 2 to Tender Dossier.
Question 68:	Item 1.6.4, “Cabinet measurements shall be at least 80 x 52 x 194 ± 2 cm.” Suggested clause “Cabinet measurements shall be at least 80± 2 cm. * 53± 2 cm. * 190± 4 cm.”
Answer 68:	Please follow Article 1.6.4 of Technical Specifications.
Question 69:	Item 1.6.5, “Upper plate of wooden plate 30 mm; Side table, partition, bottom table and covers 18 mm; Backing will be 8 mm double-faced melamine resin impregnated decor paper coated flake board. There will be four fixed compartments in the cabinet.” Suggested clause “The top surface shall be melamine coated on a 30 mm chipboard. The body shall consist of a lower table, 2 side tables and fixed shelves connected by snakes and mounted on a 8 mm chipboard with a melamine coated backing channel. Covers shall be melamine coated 18 mm chip board.”
Answer 69:	Please follow Article 1.6.5 of Technical Specifications.
Question 70:	Item 1.6.6, “Edgeband or snap-on top plate, bottom plate and covers 4 edges 1 mm, all other edges visible after mounting 1 mm PVC.” Suggested clause “The front edge of the 30 mm top table shall be 2 mm and the other edges shall be PVC tape 4. Side of bottom tables, long side of side

	table 2, front edge of fixed shelf, 4 sides of movable shelves are 0,45 mm PVC tape. 4 sides shall be 2 mm PVC tape.”
Answer 70:	Please follow Article 1.6.6 of Technical Specifications.
Question 71:	<p>Item 1.6.7, “General construction will be made with MiniFix connectors. The clips will be 128 mm in size and (spring) chrome plated. Each lid will be fixed to the shell with three bowl hinges. The height, which is made of hard plastic, is to be provided with fixed feet. One lock will be available on each of the doors. The clothes hanger pipes shall be made of aluminum mat chrome profiler to fit in the hanger ears with a diameter of 21 mm and attached to the side table and intermediate stand.”</p> <p>Suggested clause “The cupboard shall be equipped with steel minifiks. The cup shall be raised from the floor with 4 equilibrium aluminum legs. There shall be 2 flat hinges on the doors. Each door shall have a 128mm between two holes. Movable shelves, metal headed shelf pins shall be attached to it. The clothes hanger pipe connected to the side tables shall be mounted.”</p>
Answer 71:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 72:	Item 1.7, What are the dimensions of the stand and it’s place to set down? For example between the sofa, on the left/right of the sofa.
Answer 72:	Please follow Article 1.7 and subtitles of Technical Specifications.
Question 73:	<p>Item 1.7.4, “Beech wooden skeleton, wooden foot, 32 dns cutting sponge, artificial leather or fabric, rubber column will be.”</p> <p>Suggested clause “Wooden parts; it shall be produced from the ovened beech wood, in accordance with body ergonomics. Wooden records are 2.5cm x 5cm, other wooden pieces are 2-2.5 cm. thick, the product shall be made in a form suitable for the glued joining, in certain parts the joining shall be used and shall be supported by bracket parts to increase the strength.”</p>
Answer 73:	Please follow Article 1.7.4 of Technical Specifications.
Question 74:	<p>Item 1.7.5, “Seat height will be 86 ± 2 cm.”</p> <p>Suggested clause “The seat height shall be 82 ± 2 cm.”</p>
Answer 74:	Please follow Article 1.7.5 of Technical Specifications.
Question 75:	<p>Item 1.7.8, “The seating depth will be 55 ± 2 cm.”</p> <p>Suggested clause “The seating depth shall be 51 ± 2 cm.”</p>
Answer 75:	Please follow Article 1.7.8 of Technical Specifications.
Question 76:	<p>Item 1.7.10, “It shall be 47 cm above the seating surface of the upper backrest.”</p> <p>Suggested clause “it shall be 47 ± 2 cm. above the seating surface of the upper backrest.”</p>
Answer 76:	Please follow Article 1.7.10 of Technical Specifications.

Question 77:	Item 1.7.11, "Seat: 10 cm 32 dns cutting sponge - Back: 12 cm 28 dns cutting sponge - Above arm: 3 cm 28 dns cutting sponge - External parts: 1.4 cm 15 dncutting sponge with 3 cm thickness on the arm tops." Suggested clause "Sitting shall be at a density of 32 ± 2 (density, kg / m ³ at a thickness of 16 cm. The back shall be at a density of 28 ± 2 (density, kg / m ³ at a thickness of 1.5 cm. 17-21 cm thick bead fibers shall be used in the backfoot mattress. The seat cushion shall be wrapped with 300 gr of fiber."
Answer 77:	Please follow Article 1.7.11 of Technical Specifications.
Question 78:	Item 1.7.12, "The armrests will be 23 dns ± 2 cutting sponge." Suggested clause "Armrests shall be covered with artificial leather or fabric on a wooden beech carcass with 3 cm and 1.5 cm thickness 28 ± 2 (density, kg / m ³ sponge. 90 gr of fibers shall be used in the upper parts."
Answer 78:	Please follow Article 1.7.12 of Technical Specifications.
Question 79:	Item 1.7.15, "Upholstery, 10 % polyurethane, 90 % pvc and at least 530 gr / m ² . The fabric will be polyester or polyolefin and at least 220 gr / m ² , at least 40.000 will be resistant to abrasion." Suggested clause "It shall be made of fabric: 100% Polyester, Artificial Leather: 18% Polyester - 80% Pvc - 2% Polyurethane min: 400 gr / m ² max: 600 gr / m ² and it shall be possible to be cleaned with a damp cloth."
Answer 79:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 80:	Item 1.8, Study Desk + Keson, can be made with requested dimensions. But to fit Keson under the desk will not be ergonomic because of height is almost similar to study desk. You didn't ask but "L" shaped study desk will be better to use.
Answer 80:	Please follow Article 1.8 and subtitles of Technical Specifications. Please see the Corrigendum No: 2 to Tender Dossier.
Question 81:	Item 1.8.4 and 1.8.7, table height is 75 cm. top table material thickness 3 cm. There is a height of 72 cm at the bottom and the caisson height is 72 cm. Under these conditions, the caisson does not fit under the table. the height of the caisson should be at least 68 cm.
Answer 81:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 82:	Item 1.8.5, "The edges of the plates forming the top plate are covered with 2 mm pvc, the edges of the plate forming the front panel are covered with 1 mm pvc" Suggested clause "The 4 sides of the table top shall be 2 mm PVC tape."
Answer 82:	Please follow Article 1.8.5 of Technical Specifications.
Question 83:	Item 1.8.6, "A wooden screw should be used as a connection element. Feet shall have a cross-sectional dimension of 50 x 30 mm dkp 2 mm thickness,

	<p>must be electrostatic powder painted, the metal box should be made of profil. Traverse should be 40 x 40 mm, 1,5 mm dkp”</p> <p>Suggested clause “Profile legs shall be manufactured in box size 40x40x2 mm. Metal feet and intermediate inserts should be combined with aluminium comer joint elements coated with electrostatic powder paint by injection molding in special molds.”</p>
Answer 83:	Please follow Article 1.8.6 of Technical Specifications.
Question 84:	Item 1.8.6, feet cross-section is required as 50x30mm dkp. This feet dimension is so specific. 40x40 mm can be also acceptable for it? In order to increase competition, please accept 40x40 mm dkp, as well.
Answer 84:	Please follow Article 1.8.6 of Technical Specifications.
Question 85:	Item 1.8.7, drawer inside and the counter is required as metal. In the market, made from completely melamine chipboard keson is more common. In order to increase competition, accepting both metal and melamine chipboard drawer inside and the counter is highly appreciated.
Answer 85:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 86:	Item 1.8.7, we want to clarify the term "counter is made of 0,8 mm". Do you want to describe drawer glides with counter? If not, what does it mean really?
Answer 86:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 87:	Item 1.8.8, lock is center lock or it could be valid only for top drawer?
Answer 87:	Lock will be central. Please see the Corrigendum No: 2 to Tender Dossier.
Question 88:	Item 1.9.4 says 80 cm height of a cabinet, Item 1.96 asks 2 fixed compartments and 3 fixed shelves. For that kind of small cabinet, if we put 3 shelves, the files or dossiers will not fit into shelves. Nominal value should be 2 shelves. Is it an error or you really want 3 shelves?
Answer 88:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 89:	<p>Item 1.9.5, “The library top plate will be 30 mm, side plate, partition, shelf, bottom plate and covers 18 mm, backing 8 mm double face melamine resin impregnated decor paper coated chipboard. with ± 2 mm tolerance.”</p> <p>Suggested clause “The top surface shall be melamine coated ona 30 mm chipboard. The substrate shall consist of a bottom plate, two side tables and a fixed shelf, joined by saddle stitchers and mounted on a 8 mm chipboard with a melamine coated backing grooved system. Lower side and shelves shall be 18 mm. The bottom covers shall be melamine coated 18mm chip board.”</p>
Answer 89:	Please follow Article 1.9.5 of Technical Specifications.

Question 90:	<p>Item 1.9.6, “There will be 2 fixed compartments and 3 fixed shelves in the cabinet. The library top, bottom and library covers are 4 edges 1 mm, all other edges visible after mounting are 1 mm PVC with ± 2 mm tolerance.”</p> <p>Suggested clause “The front edge of the 30 mm top table is 2 mm and the other edges are PVC tape O. 4 mm of the bottom table, 2 long side of the side table, the front edge of the fixed shelf, 4 sides of the movable shelves shall be 0,40 mm PVC tape. The edge shall be 2 mm PVC tape.”</p>
Answer 90:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 91:	Item 1.9.6, “3 fixed shelves in the cabinet” is required. But, the height of the file cabinet is only 75 cm. So 3 fixed shelves can not be fitted. Only, 1 shelf can be fitted. Please check this point.
Answer 91:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 92:	<p>Item 1.9.7, “The general construction will be made with MiniFix connectors. The handle will be 128 mm in size and (spring) chrome plated. The movable shelves will be mounted on four shelf pins. The cabinet leg should be metal provided and fixed. The doors should be locked with one lock.”</p> <p>Suggested clause “The cupboard shall be mounted with steel pins. The cup shall be raised from the floor with 4 balancing aluminium feet. The doors shall have 2 flat hinges. Each lid shall have a handle with 128mm between two holes. The movable shelves shall be mounted on the metal shelf pins connected to the side tables.”</p>
Answer 92:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 93:	<p>Item 1.10, do you ask height adjustment and wheels? In the market standard, Guest chair has plastic bingo foot, not wheels. Also, position finder mechanism is used at Guest Chair, not height adjustment mechanism. So, please check and clarify these points.</p> <p>Additionally, Head support is required at 1.10.15. It is also not usual request, according to market standard. At market, Guest Chair can be supplied without head support. Please clarify “Head Support” is required or not.</p>
Answer 93:	Please follow Article 1.10 and subtitles of Technical Specifications. Please see the Corrigendum No: 2 to Tender Dossier.
Question 94:	Item 1.10.4: What is the meaning of the damortion? Does It mean shock absorber or not?
Answer 94:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 95:	<p>Item 1.10.4, “Artificial leather or fabric should be used. A 32 kg / m³ gray cut sponge should be used. Plastic armrests and star stand should be manufactured. Damortion, mechanism and plastic wheel should be used.”</p> <p>Suggested clause “It shall be made of fabric: 100% Polyester, Artificial Leather: 18% Polyester - 80% Pvc - 2% Polyurethane min: 400 gr / m² max:</p>

	600 gr / m ² and it shall be possible to be cleaned with a damp cloth. Metal Skeleton shall be on cast sponge.”
Answer 95:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 96:	Item 1.10.5, “The seating depth will be 46 ± 2 cm.” Suggested clause “The seating depth shall be 49 ± 2 cm.”
Answer 96:	Please follow Article 1.10.5 of Technical Specifications.
Question 97:	Item 1.10.6, “The seating surface width will be 48 ± 2 cm.” Suggested clause “The seating surface width will be 51 ± 2 cm.”
Answer 97:	Please follow Article 1.10.6 of Technical Specifications.
Question 98:	Item 1.10.8, “The backrest width will be 45 cm widest.” Suggested clause “The backrest width will be 48 cm widest.”
Answer 98:	Please follow Article 1.10.8 of Technical Specifications.
Question 99:	Item 1.10.10, “The sponge will be a cut sponge with a density of at least 32 kg / m ³ .” Suggested clause “The seat made of monoblock body shall be made of special aluminium molds with density of 50 ± 5 (density, kg / m ³) in accordance with body ergonomics on steel carcass.”
Answer 99:	Please follow Article 1.10.10 of Technical Specifications.
Question 100:	Item 1.10.11, “The seat belt will be manufactured from hard plastic material with injection printing.” Suggested clause “Foot group shall be made of steel pipe, the base parts shall be supported by fixed plastic lug Ø25,4 * 2 mm pipe leg, inner frame Ø16x2 mm pipe.”
Answer 100:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 101:	Item 1.10.12, “Seat foot should be produced from plastic material. Fixed plastic shoes should be used. It must have a star-like appearance” Suggested clause “Armrests shall be produced from polypropylene (PP) material and manufactured from materials that meet PAH criteria.”
Answer 101:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 102:	Item 1.10.13, “at least 2.2 mm thick should be molded. the swing must have a moving twin-spoke mechanism. Metal parts should be painted with electrostatic epoxy resin powder resistant to corrosion. The paint thickness should be at least 40 µ (microns).” Suggested clause “Monoblock injection molding sponge with an ergonomic structure of 6 cm thickness shall have a density of 50 (density, kg / m ³).”
Answer 102:	Please see the Corrigendum No: 2 to Tender Dossier.

Question 103:	<p>Item 1.10.14, “Upholstery, 2% polyurethane, 20% Polyester, 78% pvc and at least 500 gr / m2. Fabric, polyester or polyolefin will be at least 220 gr / m2 and at least 40.000 abrasion resistant. The back part will be fabric-like fabric.”</p> <p>Suggested clause “It shall be made of fabric: 100% Polyester, Artificial Leather: 18% Polyester - 80% Pvc - 2% Polyurethane min: 400 gr / m² max: 600 gr / m² and it shall be clean with a damp cloth.”</p>
Answer 103:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 104:	Item 1.10.15, we want to clarify this item is necessary or it is an error. Normally guest chairs don't have head and waist supports.
Answer 104:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 105:	Suggested clause “1.10.19 (New Clause): It shall be guaranteed for three years.”
Answer 105:	Please follow Item 32.7 Special Conditions and Article 3.6.10 of Particular Conditions under General Requirements.
Question 106:	<p>Item 1.11.5, “The edges of the plates forming the top plate are covered with 2 mm pvc, the edges of the plate forming the front panel are covered with 1 mm pvc”</p> <p>Suggested clause “Traverses and feet are connected to the table top with M6 bolts. Plastic at the point where the feet run, min. 0.5 cm - max 1.5 cm adjustable papuclar shall be connected.”</p>
Answer 106:	Please follow Article 1.11.5 of Technical Specifications.
Question 107:	<p>Item 1.11.6, “A wooden screw should be used as a connection element. Feet shall have a cross-sectional dimension of 50 x 30 mm dkp 2 mm thickness, must be electrostatic powder painted, the metal box should be made of profil. Traverse should be 40 x 40 mm, 1,5 mm dkp”</p> <p>Suggested clause “Profile legs are manufactured in box size 40x40x2 mm. The 4 sides of the tabletop shall be 2 mm PVC tape.”</p>
Answer 107:	Please follow Article 1.11.6 of Technical Specifications.
Question 108:	Suggested clause “1.11.8 (New Clause): Metal legs and intermediate inserts shall be combined with aluminum corner fittings coated with electrostatic powder paint by injection molding in special molds.”
Answer 108:	Please follow Article 1.11 and subtitles of Technical Specifications.
Question 109:	Items 1.11.16 and 1.8.6, the dimensions asked for traverse is 40x40 mm, but the sizes of feet are 50x30 mm. It is possible to produce with this dimensions but the looking will not be aesthetic. There is 10 mm difference with traverse and feet. We suggest fixing the sizes of traverse and feet.
Answer 109:	Please follow Article 1.11.16 & 1.8.6 of Technical Specifications.

Question 110:	Item 1.12 and 1.13, please specify the shape of table top. For instance, oval or rectangular.
Answer 110:	Please follow Article 1.12 & 1.13 and subtitles of Technical Specifications.
Question 111:	Item 1.12.4, "It should be at 300 x 100 x 75 cm" Suggested clause "It shall be at 300 ± 4 cm. x 140± 2 cm. x 75± 2 cm."
Answer 111:	Please follow Article 1.12.4 of Technical Specifications.
Question 112:	Item 1.13.4, "It should be at 200 x 100 x 75 cm" Suggested clause "It shall be at 200 ± 4 cm. x 140± 2 cm. x 75± 2 cm."
Answer 112:	Please follow Article 1.13.4 of Technical Specifications.
Question 113:	Item 1.14.4, "Skeleton, plastic seat and plastic back will be two separate parts. Single lever mechanism, gas shock absorber, plastic star foot, plastic wheel. 45 Kg / m ³ cast sponge, artificial leather or fabric." Suggested clause "Backrest inner plastic shall be manufactured from polypropylene material by injection in special molds according to body ergonomics. The seat pad shall be 10 ± 2 mm thick. Special aluminium molds shall be manufactured in accordance with the ergonomics of the body by giving a concave form by hot stamping process from wooden paper material."
Answer 113:	Please follow Article 1.14.4 of Technical Specifications.
Question 114:	Item 1.14.5, "The seating depth will be 42 ± 2 cm." Suggested clause "The seating depth shall be 40 ± 4 cm."
Answer 114:	Please follow Article 1.14.5 of Technical Specifications.
Question 115:	Item 1.14.6, "The seating surface width will be 39 ± 2 cm." Suggested clause "The seating surface width shall be 43 ± 2 cm."
Answer 115:	Please follow Article 1.14.6 of Technical Specifications.
Question 116:	Item 1.14.7, "Star shaped foot diameter will be 59 ± 2 cm." Suggested clause "The diameter of the star's feet of the cozmo product shall be here to be the diameter of the teller's wheel."
Answer 116:	Please follow Article 1.14.7 of Technical Specifications.
Question 117:	Item 1.14.8, "The sponge shall be a cast sponge with a density of at least 45 kg / m ³ ." Suggested clause "The seat spunk shall be manufactured in geometry suitable for body ergonomics by injecting aluminium molds with a density of 50 ± 5 (density, kg / m ³ on wooden carcass. The back sponge shall be manufactured in geometry suitable for body ergonomics by injection in aluminium molds with a density of 50 ± 5 (density, kg / m ³)"

Answer 117:	Please follow Article 1.14.8 of Technical Specifications.
Question 118:	Item 1.14.11, what does the red eye mean?
Answer 118:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 119:	Item 1.14.12, "Upholstery will be 2% polyurethane, 20% Polyester, 78% pvc and at least 500 gr / m2, Fabric Polyester or polyolefin. 220 gr / m2, at least 40.000 wear resistant." Suggested clause "It shall be made of fabric: 100% Polyester, Artificial Leather: 18% Polyester - 80% Pvc - 2% Polyurethane min: 400 gr / m2 max: 600 gr / m2 and it shall be possible to be cleaned with a damp cloth."
Answer 119:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 120:	Item 1.15, we need to simple photo or plotting of hall stand.
Answer 120:	Please follow Article 1.15 and subtitles of Technical Specifications.
Question 121:	Item 1.17, please specify the material of mirror frame.
Answer 121:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 122:	Item 1.18, wish & complaint box is 31x19 ± 2. But what is its depth?
Answer 122:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 123:	Item 1.18.1, It will be manufactured from high alloy carbon steel sheet material. The use of DKP sheet in these products is more common. The places of use of carbon products are preferred for curing precious metals or for making a mold or for glass products. Our request "It will be manufactured from high alloy carbon steel sheet material or DKP sheet material."
Answer 123:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 124:	Item 1.19.4, "The extinguisher of the fire extinguisher will be dry chemical powder." Suggested clause "The body of the device shall be manufactured by 1.5 mm 7114 quality Ereğli Erdemir Iron Steel plant and the lower cover shall be made of 2 mm. 7114 quality plaster."
Answer 124:	Please follow Article 1.19.4 of Technical Specifications.
Question 125:	Item 1.19.5, "Extinguisher weight will be 6 kg." Suggested clause "The dusts used shall belong to the product manufactured and shall be presented with an import certificate. 6 Kg net Abc Map rate should be filled with 40% Furex German extinguishing material wit TSE certificate."
Answer 125:	Please follow Article 1.19.5 of Technical Specifications.

Question 126:	Item 1.19.6, "The pusher type will be internally pressurized." Suggested clause "The equipment shall not be exposed to wet ground, the devices shall not be subjected to corrosion and special plastic feet shall be used."
Answer 126:	Please follow Article 1.19.6 of Technical Specifications.
Question 127:	Item 1.19.7, "The test pressure will be 25 kg / cm ² ." Suggested clause "The equipment shall not be exposed to wet ground, the devices shall not be subjected to corrosion and special plastic feet shall be used."
Answer 127:	Please follow Article 1.19.7 of Technical Specifications.
Question 128:	Item 1.19.8, "The working pressure will be 17 kg / cm ² ." Suggested clause "The hoses on the device which are used to interfere with the fire with water shall be flood-proof and shall be TSE certified."
Answer 128:	Please follow Article 1.19.8 of Technical Specifications.
Question 129:	Item 1.19.9, "The discharge time will be 9 seconds." Suggested clause "The reliefvalve located on the device shall be CE certified. In addition, there shall be a special valve system and a safety system to ensure that the manometer shall be correctly displayed."
Answer 129:	Please follow Article 1.19.9 of Technical Specifications.
Question 130:	Item 1.19.10, "Spray distance will be max. 4 – 5 metres" Suggested clause "The devices must be electrostatic painted"
Answer 130:	Please follow Article 1.19.10 of Technical Specifications.
Question 131:	Item 1.19.11, "Exhalant additive will be less than 8%." Suggested clause "6 Kg Abc dust belonging to fire extinguishing devices shall be taken from abroad accredited laboratory and have performance test."
Answer 131:	Please follow Article 1.19.11 of Technical Specifications.
Question 132:	Item 1.19.12, "Hair thickness will be 2 mm Iron Plated Press." Suggested clause "The boom shall consist of explosion pressure 100 bar, working pressure 18 bar."
Answer 132:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 133:	Item 3.0.9, In the document c4f_annexIItechspec, under LOT 3, item 3.0.9, asking the "After Sales Service Qualification Certificate". As a regulation dated 24.04.2011 and No: 27914 published in official gazetta it is clearly stated that, TC. Ministry of Customs and Trade will not issue any certificate for surgical instruments. You can check all products which are needed to get "After Sales Service Qualification Certificate" as a annex of this

	regulation. So we kindly ask you to modify or remove item LOT3, 3.0.9 from Technical Specifications, because manufacturers of surgical instruments are not allowed to get mentioned certificate from TC. Ministry of Customs and Trade since 2011.
Answer 133:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 134:	Item 3.09, "T.C. Ministry of Industry and Trade After Sales Service Qualification Certificate" is required from domestic, importer and participating company. However, Surgical Instrument are outside the scope of "After Sales Service Guide" (Resmi Gazete Tarihi: 13.06.2014 Resmi Gazete Sayısı: 29029). So, "T.C. Ministry of Industry and Trade After Sales Service Qualification Certificate" should not be required. Please cancel item 3.09 from general conditions for surgical instrument at LOT-3.
Answer 134:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 135:	Item 3.1 to 3.9, Items 3.1, 3.2, 3.3, 3.4, 3.5, 3.6 and A- Laryngoscope Set of item 3.9 are defined as "Surgical Instruments" at all over the world. In order to supply quality solution and protect the end users, we kindly recommend brand integrity at these items, for benefit of the contractor and end users. These means, the participator should provide single brand at item 3.1, 3.2, 3.3, 3.4, 3.5, 3.6 and A- Laryngoscope Set of item 3.9.
Answer 135:	Please follow Article 3.1 to 3.9 of Technical Specifications. Please see the Corrigendum No: 2 to Tender Dossier.
Question 136:	Item 3.1.9, Small Surgery Set, "Stapler" is required. In the market, Stapler is supplied as Stapler and Stapler Removing Forceps with together. In order to eliminate any misunderstanding, please clarify this point.
Answer 136:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 137:	Item 3.2, Inspection set, there is repetition after 3.2.18. Between 3.2.19 and 3.2.30 should be canceled. Please clarify this point.
Answer 137:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 138:	Item 3.2.17 and 3.2.18, At item 3.2.17 and 3.2.18, same penset are required. We thought that 3.2.18 should be "Tissue Penset", please clarify it.
Answer 138:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 139:	Item 3.8.5, "The negatoscope shall have an aesthetic appearance, the device measurements shall not exceed 460 x 520 x 25 mm." Requested amendment "Device measurement as 460 x 520 x 100 mm ±100"
Answer 139:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 140:	Item 3.8.5, "The negatoscope shall have an aesthetic appearance, the device measurements shall not exceed 460 x 520 x 25 mm."

	Requested amendment “The negatoscope shall have an aesthetic appearance, the device measurements shall not exceed 460 x 520 x 25 mm (\pm 30 mm).”
Answer 140:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 141:	Item 3.8.8, “The weight of the device shall be a maximum 5 kg.” Requested amendment “weight of the device as 5 kg \pm 1kg”
Answer 141:	Please follow Article 3.8.8 of Technical Specifications.
Question 142:	Item 3.8.10, “The device shall operate between 220-240 volts without any problems. The input power of the device shall be 12 volt 1.66 mA.” Requested amendment “The device shall operate between 220-240 volts without any problems. The input power of the device shall be 12 volt 1.66 mA or 220 Volt”
Answer 142:	Please follow Article 3.8.10 of Technical Specifications.
Question 143:	Item 3.8.13, “The stainless steel body of the device shall be made from a single material and be paintable.” Requested amendment “The stainless steel body of the device shall be made from a single material or shall be made by DKP sheet and be paintable or already painted”
Answer 143:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 144:	Item 3.8.13, “The stainless steel body of the device shall be made from a single material and be paintable.” Requested amendment “The body of the device shall be made from corrosion resistant steel material.”
Answer 144:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 145:	Item 3.9, please specify type of the blade, Macintosh or Miller. Please note that Macintosh is most common type. In order to take same offer from all participator, please clarify the type of blade.
Answer 145:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 146:	Item 3.9, CPAP set is required in emergency set. In world there are two brands of such product one is Malaysian made and other one is made in USA. These both origins are not available in terms and conditions document. Can we still submit our offer with American product?
Answer 146:	Please follow Article 3.9 and subtitles of Technical Specifications. Please see the Corrigendum No: 2 to Tender Dossier.
Question 147:	Item 3.9.7, B - Oxygen Tube-attachable Respiration Set This item needs clarification. Possible photo of the item may be helpful.
Answer 147:	Please follow Article 3.9.7 of Technical Specifications.

Question 148:	Item 3.9.8.7, “Child bag volume 5500 ml, pressing volume 300 ml and reservoir volume shall be 2500 ml.” According to the required specifications of the needed ambu device, the child bag volume should be 500 ml’s. The number of distribution of Adult and Children should be stated.
Answer 148:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 149:	Item 4.3.10, “The cuff shall be made of very sturdy fabric and fiber with a 2 (two) phase hook on the cuff” According to technical requirements cuff of blood pressure machine must have hooks. We request to change the condition as “The cuff shall be made of very sturdy fabric and fiber with a 2 phase hook on the cuff OR self adhesive”
Answer 149:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 150:	Item 4.3.16, “The dimensions of the cuff shall be 10 cm x 47 (± 0,5) cm” In order to submit our offer we need the dimensions of cuff as 13 x 42 cm
Answer 150:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 151:	Item 4.3.16, “The dimensions of the cuff shall be 10 cm x 47 (± 0,5) cm” 47 cm is the dimension for Adult Cuff. The general dimension for required 48 mm manometer is 39 cm. the dimension of the cuff should be changed as 39 cm.
Answer 151:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 152:	Item 4.4.15, “Hooked outer cuff dimensions: 49,50 x 14,50 cm, inner bladder dimensions: 22 x 12 cm.” The current stated outer cuff dimensions are not produced by any firm. If there is a production for these numbers, it only states one firm and it prevents competition. Dimension for the adult outer cuff should be changed as 47 cm to 13 cm by putting ± 2.5 cm.
Answer 152:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 153:	Item 4.4.21, “Adhesive children’s outer cuff dimensions: 37 x 10.50 cm Inner bladder dimensions 18 x 9 cm.” Name of the item is Adult Pressure Blood Monitor. In 4.4.21 children’s cuff size is stated, so this article should be removed from the technical specifications.
Answer 153:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 154:	Item 4.6.6, “Body temperature measurement range should be between 34 °C - 42,5 °C.” In order to submit our offer we need to revize maximum temperature capacity of thermometer as 42.2

Answer 154:	Please follow Article 4.6.6 of Technical Specifications.
Question 155:	Item 4.7.9, “In case of contamination or dust, the product shall contain an otoscope head, one ophthalmoscope head and reusable 2.5, 3.5, 4.5, 5, 9 mm reusable speculum in a hard plastic case” We request to change the size of reusable speculum from 2.5, 3.5 and 4.5 mm to 2,3 and 4 mm
Answer 155:	Please follow Article 4.7.9 of Technical Specifications.
Question 156:	Item 4.9.14, “The devices collection jar shall be made of sturdy plastic material and carried with a single hand with the handle on its lid.” Requested amendment “The devices collection jar shall be made of sturdy plastic material and carried with a single hand with the handle on its lid or on its body.” <i>(Handle might break the jar lid, the jar is more durable if handle is placed on the body of the jar.)</i>
Answer 156:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 157:	Item 4.9.17, “The bracket which enables the waste collection jars to be attached to the aspirators, shall be found on the jars lid. The bracket shall be found on the jars lid.” Requested amendment “The bracket which enables the waste collection jars to be attached to the aspirators, shall be found on the jars lid or on its body. The bracket shall be found on the jars lid or on the body of the jar.” <i>(Handle might break the jar lid, the jar is more durable if handle is placed on the body of the jar.)</i>
Answer 157:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 158:	Item 4.9.24, “The back of the device shall have pneumatic control foot pedal connection port.” Requested amendment “The back of the device shall have pneumatic control foot pedal connection port or electric control foot pedal.”
Answer 158:	Please follow Article 4.9.24 of Technical Specifications.
Question 159:	Item 4.9.33, Materials to be given together with the device <ul style="list-style-type: none"> • 2, 5 liter jars. • 2 jar carriers/fittings, • 2 jar lids - with hydrophobic filters • 2 hydrophobic filters • 1 Yankauer aspiration set • 1 cannula storage container will be provided Requested amendment “At least 2,5 liter jars”

Answer 159:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 160:	Item 4.10.14, "The color temperature of the examination lamp shall be 6500 ± 250 Kelvin." Requested amendment "The color temperature of the examination lamp shall be between 3.500 - 5.000 Kelvin."
Answer 160:	Please follow Article 4.10.14 of Technical Specifications.
Question 161:	Item 4.10.14, "The color temperature of the examination lamp shall be 6500 ± 250 Kelvin." Requested amendment "The color temperature of the examination lamp shall be 4000- 6500 ± 250 Kelvin."
Answer 161:	Please follow Article 4.10.14 of Technical Specifications.
Question 162:	Item 4.10.17, "Static paint shall be used in the spiral lever material." Requested amendment "Static paint shall be used in the spiral lever material or PVC coated overchrome"
Answer 162:	Please follow Article 4.10.17 of Technical Specifications.
Question 163:	Item 4.11.7, "Shall have a usable volume of 24 lt. (in order of width x height x depth) and 29 x 31 x 27 (± 2) cm. internal measurements, 58 x 50 x 41 (± 2) cm will be in external measurements." According to the dimensions, the minimum usable volume should be 20 liters (2.7x2.9x2.5=19,57 liters). But, the usable volume was indicated as 24 liters. The usable volume can be amended as; "Shall have a usable volume of minimum 20 lt. (in order of width x height x depth) and 29 x 31 x 27 (± 2) cm. internal measurements, 58 x 50 x 41 (± 2) cm will be in external measurements."
Answer 163:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 164:	Item 4.11.26, "The device shall continue to operate if it is within the set alarm limits when the device is switched off during the operation after the programmed constant temperature process, if it is outside the set limits, it shall stop the process and signal the power off. It shall prompt the user whether to continue or finish the process." It is risky to leave the decision to continue or finish the process when the temperature is out of alarm limits. In order to favour competition, we request that the sentence shall be amended as: "In case of a power failure, the device shall continue to operate if it is within the set alarm limits. Otherwise, the program should be cancelled automatically."
Answer 164:	Please follow Article 4.11.26 of Technical Specifications.

Question 165:	<p>Item 4.11.27, “The program can be scheduled to start at any time between 1 minute and 99 hours and 59 minutes.”</p> <p>In order to favour competition, we request that the sentence shall be amended as:</p> <p>“The program can be scheduled to start at any time between 1 minute and 99 hours”</p>
Answer 165:	Please follow Article 4.11.27 of Technical Specifications.
Question 166:	<p>Item 4.11.28, “The device shall have a timer that can be set between 1 - 99 hours and 59 minutes in 1 minute increments.”</p> <p>Set time timer count backwards after reaching the set temperature, the process will start and stop heating at the end of the long process will give a signal. In order to favour competition, we request that the sentence shall be amended as:</p> <p>“The device shall have a timer that can be set between 1 - 99 hours in 1 minute increments.”</p>
Answer 166:	Please follow Article 4.11.28 of Technical Specifications.
Question 167:	<p>Item 4.11.31, “Thermal insulation of the device shall be made of glass fiber and electrical insulation shall be in accordance with TS 2000.”</p> <p>Rock wool is also commonly used by manufacturers as an insulation material in such devices. In order to favour competition, we request that the sentence shall be amended as:</p> <p>“Thermal insulation of the device shall be made of glass fiber or rock wool and electrical insulation shall be in accordance with TS 2000.”</p>
Answer 167:	Please follow Article 4.11.31 of Technical Specifications.
Question 168:	<p>Item 4.11.35, “The device shall be packaged in a sturdy carton box that is suitable for transport, wrapped with air cushion nylon.”</p> <p>Our devices are packed to conform to international transportation conditions. Therefore the internal support material in the package cannot be accepted as a differentiating factor and may vary from manufacturer to manufacturer. Therefore, in order to favour competition, we request that the sentence shall be amended as:</p> <p>“The packaging shall be suitable for transportation and handling”</p>
Answer 168:	Please follow Article 4.11.35 of Technical Specifications.
Question 169:	<p>Item 4.11.39, “Manufacturers EN ISO 9001: 2008 in accordance with ISO 13485: 2013 Quality Management System.”</p> <p>The certification process is performed in accordance with ISO 13485:2012 until 2019. In order to favour competition, we request that this standard is amended as</p> <p>“Manufacturers EN ISO 9001: 2008 in accordance with ISO 13485: 2012 Quality Management System.”</p>

Answer 1689:	Please follow Article 4.11.39 of Technical Specifications.
Question 170:	Item 5.1.8, "At least 20 devices shall be given with the strips." 20 monitors are asked to supply with 1 pack of strips. It needs a correction as 20 packs of strips with one monitor.
Answer 170:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 171:	Item 5.1.10, "The device to be given with the strips shall be able to calibrate automatically." We want to completely remove 5.1.10.
Answer 171:	Please follow Article 5.1.10 of Technical Specifications.
Question 172:	Item 5.1.11, "The company shall do a demo and receive approval before the tender." A Demo is required before the tender. We want to reconfirm this.
Answer 172:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 173:	Item 5.1.13, "The measurement strip shall be able to measure at the 20-70% hematocrit range." We want to change the Hematocrit range of Glucometer from %20-70 to %20-60.
Answer 173:	Please follow Article 5.1.13 of Technical Specifications.
Question 174:	Item 5.1.14, "Each strip shall be covered in aluminum foil and shall not be affected from moisture, heat and light." According to 5.1.14 each strip must be covered in Aluminium foil. Such strip is not available in market we request to remove condition 5.1.14 completely.
Answer 174:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 175:	Item 5.1.14, "Each strip shall be covered in aluminum foil and shall not be affected from moisture, heat and light." The specifications that is mentioned at the above is just available for Abbott brand, so we kindly request the removal of this mentioned Article because fair competition is not possible because of the opportunity to offer a device for a certain brand and equal opportunity for all tenderers.
Answer 175:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 176:	Item 5.2, Origin requirements are not met in Turkey for measuring tape. We request to permit Chinese origin for his product.
Answer 176:	Please follow Article 5.2 of Technical Specifications.
Question 177:	Item 5.3.6, "Electrical characteristics of the device: 220V/50Hz - not more than 0.4 Amp/70 Watt."

	Suggested clause “Electrical characteristics of the device: 220V/50Hz - not more than 0.5 Amp/70 Watt.”
Answer 177:	Please follow Article 5.3.6 of Technical Specifications.
Question 178:	Item 5.3.7, “The devices compressor air flow in a minute shall be at least 6.5 liters and the outlet pressure must be at least 3.0 bar.” Suggested clause “The devices compressor air flow in a minute shall be at least 6.5 liters and the outlet pressure must be at least 2.41 bar.”
Answer 178:	Please follow Article 5.3.7 of Technical Specifications.
Question 179:	Item 5.3.11, “The aerosol particles formed by the device shall have a diameter (MMAD) of 0.30 micron to 6 microns. Atomizer capacity shall be 6 ml” Suggested clause “The aerosol particles formed by the device shall have a diameter (MMAD) of 0.50 micron to 10 microns. Atomizer capacity shall be 5 ml”
Answer 179:	Please follow Article 5.3.11 of Technical Specifications.
Question 180:	Item 5.4.4, “The weighing capacity of the device shall be 20 kg (+5)” Suggested clause “The weighing capacity of the device shall be 30 kg”
Answer 180:	Please follow Article 5.4.4 of Technical Specifications.
Question 181:	Item 5.4.5, “The device shall have a sensitivity of 5 gr up to 10 kg and 15 gr over 10 kg.” Suggested clause “The device shall have a sensitivity of 5 gr up to 10 kg and 10 gr over 10 kg.”
Answer 181:	Please follow Article 5.4.5 of Technical Specifications.
Question 182:	Item 5.4.7, “The wide digital indicator shall be easy-to-touch.” Suggested clause “It must be keyed”
Answer 182:	Please follow Article 5.4.7 of Technical Specifications.
Question 183:	Item 5.4.8, “The screen of the weighing tool shall be mounted on the device.” Suggested clause “Size measuring bar may not be mounted or not mounted on the device.”
Answer 183:	Please follow Article 5.4.8 of Technical Specifications.
Question 184:	Item 5.4.13, “The scale pan and body shall be of a durable structure.” We request to revise the condition as The scale pan and body shall be of a durable structure. “Length measuring scale can be provided separately.”
Answer 184:	Please follow Article 5.4.13 of Technical Specifications.
Question 185:	Item 5.5.4 “The scale shall be able to weigh 1 kg to 150 kg.”

	The firm request that the scale shall be able to weigh 1 kg to 200 kg.
Answer 185:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 186:	Item 5.7.7, “The measuring range of the device shall be 20-300 bpm and the measurement accuracy shall be ± 1 .” We request to change BPM 20-300 to BPM 0-254
Answer 186:	Please follow Article 5.7.7 of Technical Specifications.
Question 187:	Item 5.7.7, “The measuring range of the device shall be 20-300 bpm and the measurement accuracy shall be ± 1 .” Suggested clause “The measuring range of the device shall be 30-235 bpm and the measurement accuracy shall be ± 2 .”
Answer 187:	Please follow Article 5.7.7 of Technical Specifications.
Question 188:	Item 5.7.8, “The saturation measurement range of the device shall be between 0-100% and the measurement accuracy shall be ± 2 between 70% - 100%.” Suggested clause “The saturation measurement range of the device shall be between 0-100% and the measurement accuracy shall be ± 2 between 70% - 99%.”
Answer 188:	Please follow Article 5.7.8 of Technical Specifications.
Question 189:	Item 5.7.9, “The device shall be able to operate continuously for at least 50 hours with 2 AAA alkaline batteries.” Suggested clause “The device shall be able to operate continuously for at least 30 hours with 2 AAA alkaline batteries.”
Answer 189:	Please follow Article 5.7.9 of Technical Specifications.
Question 190:	Item 5.7.10, “The dimensions of the device shall be 57 mm x 35 mm x 27 mm.” Suggested clause “The dimensions of the device shall be 58 ± 2 mm x 32 ± 2 mm x 34 ± 2 mm.”
Answer 190:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 191:	Item 5.7.10, “The dimensions of the device shall be 57 mm x 35 mm x 27 mm.” We want to change the dimensions of device as 61 x 36 x 32 mm.
Answer 191:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 192:	Item 5.7.11, “The weight of the device shall not exceed 50 grams including the battery.” We request to change the maximum weight from 50 to 58 grams.
Answer 192:	Please see the Corrigendum No: 1 to Tender Dossier

Question 193:	Item 5.7.13, “The device shall be a large and easily readable color OLED display.” Suggested clause “The device shall be a large and easily readable color LED display.”
Answer 193:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 194:	Item 5.7.17, “The device shall be able to operate at -20-50 °C.” Suggested clause “The device shall be able to operate at -10-40 °C.”
Answer 194:	Please follow Article 5.7.17 of Technical Specifications.
Question 195:	Item 5.8, Do you demand Binocular Microscope as infinity-corrected optics according to the LOT 5 Article 5.8?
Answer 195:	Please follow Article 5.8 of Technical Specifications.
Question 196:	Item 5.9.4, “The device shall be portable and shall have a 2 MHz probe with a monoblock body and a minimum diameter of 20 mm.” Suggested clause “The device shall have a portable 2 MHz probe.”
Answer 196:	Please follow Article 5.9.4 of Technical Specifications.
Question 197:	Item 5.9.8, “The digital display of the device shall have a sign indicating whether the device is in the right place.” We request to completely remove this condition.
Answer 197:	Please follow Article 5.9.8 of Technical Specifications.
Question 198:	Item 5.9.16, “If the device does not receive a signal, it shall automatically shut off after 3 minutes.” Suggested clause “If no signal shall be received from the device, it shall turn off automatically after 5 minutes.”
Answer 198:	Please follow Article 5.9.16 of Technical Specifications.
Question 199:	Item 6.1.4, “The defibrillator device will be the type with the monitor and it will be capable of being used in both the automatic external and manual defibrillation modes.” Suggested clause “The defibrillator device will be the type with the monitor and it will be capable of being used in manual defibrillation modes.”
Answer 199:	Please follow Article 6.1.4 of Technical Specifications.
Question 200:	Item 6.1.6, “The device will be suitable for pediatric / child and adult use. The device will have defibrillation spoons for adult and pediatric / child use.” Suggested clause “The device will be suitable for pediatric / child and adult use. The device will have defibrillation for adult and pediatric / child use.”
Answer 200:	Please follow Article 6.1.6 of Technical Specifications.

Question 201:	Item 6.1.8, “At least 3 waveforms (for ECG, SPO2, etc. waveforms) on the screen should be viewed at the same time.” Suggested clause “The device should display ECG waveforms on the screen”
Answer 201:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 202:	Item 6.1.8 “At least 3 waveforms (for ECG, SPO2, etc. waveforms) on the screen should be viewed at the same time.” 1. Request: At least 2 waveforms (for ECG, SPO2, etc. waveforms) on the screen should be viewed at the same time. 2. Request: At least 3 parameters (for ECG, SPO2, etc. waveforms) on the screen should be viewed at the same time.
Answer 202:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 203:	Item 6.1.9, “The maximum defibrillation energy of the device will be 270 joules.” Suggested clause “The maximum defibrillation energy of the device will be at least 270 joules”
Answer 203:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 204:	Item 6.1.9, “The maximum defibrillation energy of the device will be 270 joules.” Suggested clause “The maximum defibrillation energy of the device will be 300 joules.”
Answer 204:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 205:	Item 6.1.9, “The maximum defibrillation energy of the device will be 270 joules.” As per as “European Resuscitation Council Guidelines for Resuscitation 2015” guidelines: https://ercguidelines.elsevierresource.com/european-resuscitation-council-guidelines-resuscitation-2015-section-3-adult-advanced-life-support/fulltext therefore, we need amendment as following: The minimum defibrillation energy of the device will be 200 joules. Also the device should make the defibrillation as biphasic as described as <ul style="list-style-type: none"> • Smartbiphasic • Actybiphasic • Rectilinear • 360 Joules BTE • Multi pulse Biowave
Answer 205:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 206:	Item 6.1.10, “The device will be able to defibrillate at least 100 times with at least 150 joules when the battery is full.”

	Suggested clause “The device will be able to defibrillate at least 100 times with full energy (300 Joule) from a new fully charged battery”
Answer 206:	Please follow Article 6.1.10 of Technical Specifications.
Question 207:	Item 6.1.11, “The device will be able to monitor for at least 3 hours with fully charged battery.” Suggested clause “The device will be able to monitor for at least 6 hours with fully charged battery.”
Answer 207:	Please follow Article 6.1.11 of Technical Specifications.
Question 208:	Item 6.1.12, “The device will reach maximum energy level in maximum 5 seconds with fully charged battery for manual defibrillation.” Suggested clause “The device will reach maximum energy level in maximum 8 seconds with fully charged battery for manual defibrillation.”
Answer 208:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 209:	Item 6.1.12, “The device will reach maximum energy level in maximum 5 seconds with fully charged battery for manual defibrillation.” Suggested clause, “The device will reach maximum energy level in no more than maximum 8 seconds with fully charged battery for manual defibrillation”
Answer 209:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 210:	Item 6.1.13, “After defibrillation, the maximum recharge time must be 3 seconds.” Suggested clause “The device will reach maximum energy level in maximum 8 seconds with fully charged battery for manual defibrillation.”
Answer 210:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 211:	Item 6.1.13, “After defibrillation, the maximum recharge time must be 3 seconds.” Suggested clause “After defibrillation, the maximum recharge time must be at most 8 seconds.”
Answer 211:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 212:	Item 6.1.14, “The defibrillation energy of the device will be gradually adjustable.” Suggested clause “The defibrillation energy of the device will be gradually adjustable: 1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 20, 30, 50, 70, 100, 150, 200, 250, 300 Joules.”
Answer 212:	Please follow Article 6.1.14 of Technical Specifications.
Question 213:	Item 6.1.16, “The device will provide biphasic defibrillation. Defibrillation process will be done with disposable pads and spoons. Spoons and pads will be suitable for adult and pediatric / child use. The device will be

	<p>accompanied by sets of adult and pediatric / children disposable pads, 2 for each.”</p> <p>Suggested clause “The device will provide biphasic defibrillation. Defibrillation process will be done with 2 paddles. The paddles shall be suitable for adult and paediatric use. The adult paddle shall retractable to transform the paddles to paediatric paddles.”</p>
Answer 213:	Please follow Article 6.1.16 of Technical Specifications.
Question 214:	<p>Item 6.1.17, “The instrument will be able to measure EtCO2 (End-Tidal Carbon Dioxide), SPO2 and NIBP with at least 12 channels ECG recording.”</p> <p>Suggested clause “The instrument will be able to measure at least 3 channels ECG recording.”</p>
Answer 214:	Please follow Article 6.1.17 of Technical Specifications.
Question 215:	<p>Item 6.1.17, “The instrument will be able to measure EtCO2 (End-Tidal Carbon Dioxide), SPO2 and NIBP with at least 12 channels ECG recording.”</p> <p>As per as “European Resuscitation Council Guidelines for Resuscitation 2015” guidelines: https://ercguidelines.elsevierresource.com/european-resuscitation-council-guidelines-resuscitation-2015-section-3-adult-advanced-life-support/fulltext</p> <p>The purpose of ECG monitoring is not for as an ECG machine with 12 leads or with 12 channels. 3 or 5 leads ECG monitoring in 2 channels can meet all defibrillation requirements and also monitor other parameters in your requested min. 6.5 inches display. The project is buying dedicated resting ECG machines separately. Regarding ECG during defibrillation,</p> <p>“Whenever a diagnosis of asystole is made, check the ECG carefully for the presence of P waves, because this may respond to cardiac pacing. There is no benefit in attempting to pace true asystole. In addition, if there is doubt about whether the rhythm is asystole or extremely fine VF, do not attempt defibrillation; instead, continue chest compressions and ventilation. Continuing high-quality CPR however may improve the amplitude and frequency of the VF and improve the chance of successful defibrillation to a perfusing rhythm”. Therefore, we need amendment as following:</p> <p>“The instrument will be able to measure EtCO2 (End-Tidal Carbon Dioxide), SPO2 and NIBP with at least 2 channels ECG recording with 3 or 5 leads.”</p>
Answer 215:	Please follow Article 6.1.17 of Technical Specifications.
Question 216:	<p>Item 6.1.18, “The device should be capable of ECG monitoring with ECG cables and pads. With ECG cable, leads I, II, III, AVR, AVL, AVF, V1 - V6 will be displayed on the screen and the desired lead can be selected.”</p> <p>Requested amendment “The device should be capable of ECG monitoring with ECG cables and pads. With ECG cable, leads I, II, III, AVR, AVL,</p>

	AVF, V1 - V6 will be displayed and at least one select ECG derivation can be viewed single channel on screen.”
Answer 216:	Please follow Article 6.1.18 of Technical Specifications.
Question 217:	<p>Item 6.1.18, “The device should be capable of ECG monitoring with ECG cables and pads. With ECG cable, leads I, II, III, AVR, AVL, AVF, V1 - V6 will be displayed on the screen and the desired lead can be selected.”</p> <p>As per as “European Resuscitation Council Guidelines for Resuscitation 2015” guidelines: https://ercguidelines.elsevierresource.com/european-resuscitation-council-guidelines-resuscitation-2015-section-3-adult-advanced-life-support/fulltext</p> <p>The purpose of ECG monitoring is not for as an ECG machine with 12 leads or with 12 channels. 3 or 5 leads ECG monitoring in 2 channels can meet all defibrillation requirements and also monitor other parameters in your requested min. 6.5 inches display. The project is buying dedicated resting ECG machines separately. Regarding ECG during defibrillation,</p> <p>“Whenever a diagnosis of asystole is made, check the ECG carefully for the presence of P waves, because this may respond to cardiac pacing. There is no benefit in attempting to pace true asystole. In addition, if there is doubt about whether the rhythm is asystole or extremely fine VF, do not attempt defibrillation; instead, continue chest compressions and ventilation. Continuing high-quality CPR however may improve the amplitude and frequency of the VF and improve the chance of successful defibrillation to a perfusing rhythm”. Therefore, we need amendment as following:</p> <p>“The device should be capable of ECG monitoring with ECG cables and pads. With 5 lead ECG cable, leads I, II, III, AVR, AVL, AVF, V will be displayed on the screen and the desired lead can be selected.”</p>
Answer 217:	Please follow Article 6.1.18 of Technical Specifications.
Question 218:	<p>Item 6.1.18, “The device should be capable of ECG monitoring with ECG cables and pads. With ECG cable, leads I, II, III, AVR, AVL, AVF, V1 - V6 will be displayed on the screen and the desired lead can be selected.”</p> <p>Suggested clause “The device should be capable of ECG monitoring with ECG cables and pads: Paddle, I, II, III V, aVR, aVL, aVF (only with 5L cable)”</p>
Answer 218:	Please follow Article 6.1.18 of Technical Specifications.
Question 219:	<p>Item 6.1.19, “Heart rate measurement range to be measured by ECG feature WILL BE at least 15 to 300 beats per minute.”</p> <p>Suggested clause “Heart rate measurement range to be measured by ECG feature will be 30 - 250 bpm with an accuracy of ±10% or ±5 bpm”</p>
Answer 219:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 220:	Item 6.1.19, “Heart rate measurement range to be measured by ECG feature WILL BE at least 15 to 300 beats per minute.”

	Suggested clause “Heart rate measurement range to be measured by ECG feature will be at least 30 to 300 beats per minute.”
Answer 220:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 221:	Item 6.1.20, “The device will measure oxygen saturation (SPO2) as standard. The SPO2 measurement range will be at least 70 - 100%. All accessories required for measurement (sensor, cable, etc.) will be supplied with the device.” Suggested clause “Removal from Technical Specification”
Answer 221:	Please follow Article 6.1.20 of Technical Specifications.
Question 222:	Item 6.1.21, “The device will perform NIBP (Non-invasive blood pressure) measurement as standard. Systolic and diastolic pressure values should be available. Systolic pressure can be measured at a range of at least 60 - 250 mmHg, diastolic pressure at least 40 - 200 mmHg.” Suggested clause “Removal from Technical Specification”
Answer 222:	Please follow Article 6.1.21 of Technical Specifications.
Question 223:	Item 6.1.22, “The device will measure EtCO2 (End-Tidal Carbon Dioxide) as standard. The measurement range shall be at least 0-99 mmHg. All accessories required for measurement (sensor, cable, etc.) will be supplied with the device.” Suggested clause “Removal from Technical Specification”
Answer 223:	Please follow Article 6.1.22 of Technical Specifications.
Question 224:	Item 6.1.23, “The device will operate in manual defibrillation and automatic modes (AED).” Suggested clause “The device will operate in manual defibrillation”
Answer 224:	Please follow Article 6.1.23 of Technical Specifications.
Question 225:	Item 6.1.25, “The printer speed will be 25 mm / s and 50 mm / s.” Suggested clause “The printer speed will be 25 mm / s”
Answer 225:	Please follow Article 6.1.25 of Technical Specifications.
Question 226:	Item 6.1.26, “The device will have memory for at least 500 events, or memory card (CF card, memory card etc.) or memory for at least 360 minutes of continuous ECG data or at least 1 GB of external memory.” Suggested clause “The device will have memory for at least 10 events”
Answer 226:	Please follow Article 6.1.26 of Technical Specifications.
Question 227:	Item 6.1.26, “The device will have memory for at least 500 events, or memory card (CF card, memory card etc.) or memory for at least 360 minutes of continuous ECG data or at least 1 GB of external memory.”

	<p>As per as “European Resuscitation Council Guidelines for Resuscitation 2015” guidelines: no such non-clinical needs are described. Current specs. describes specific product with their used technical features. We need amendment as following:</p> <p>“The device will have memory for at least 500 events, or memory card (CF card, memory card, USB etc.) or memory for at least 360 minutes of continuous ECG data or at least 1 GB of external memory.”</p>
Answer 227:	Please follow Article 6.1.26 of Technical Specifications.
Question 228:	<p>Item 6.1.26, “The device will have memory for at least 500 events, or memory card (CF card, memory card etc.) or memory for at least 360 minutes of continuous ECG data or at least 1 GB of external memory.”</p> <p>Requested amendment “The device will have memory for at least 500 events, or memory card (CF card, memory card etc.) or memory for at least 360 minutes of continuous ECG data or at least 1 GB of external memory or device should have internal memory for minimum 30 defibrillation event”</p>
Answer 228:	Please follow Article 6.1.26 of Technical Specifications.
Question 229:	<p>Item 6.1.27, “The saved data will be transferred to the computer. If any software and / or hardware is required to transfer and display the data on a computer, such software and / or hardware shall be provided free of charge with the consignment.”</p> <p>Requested amendment “The saved data will be transferred to the computer. If any software and / or hardware is required to transfer and display the data on a computer, such software and / or hardware shall be provided free of charge with the consignment. If saved data can be called on device screen and be examined, not need to data transfer between device and computer.”</p>
Answer 229:	Please follow Article 6.1.27 of Technical Specifications.
Question 230:	<p>Item 6.1.29, “The battery will be fully charged within 5 hours at most.”</p> <p>Suggested clause “The battery will be fully charged within 6 hours at most.”</p>
Answer 230:	Please follow Article 6.1.29 of Technical Specifications.
Question 231:	<p>Item 6.1.31, “The batteries can also be charged with the 12 V charging voltage via the device or wall mount kit.”</p> <p>Requested amendment “The batteries can also be charged with the 12 V charging voltage via the device or wall mount kit. If not possible, supplier will provide inverter for 12V charging with each device.”</p>
Answer 231:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 232:	<p>Item 6.1.35, “The device's IP (Ingress Protection) class will be at least IP-32.”</p> <p>Suggested clause “The device's IP (Ingress Protection) class will be at least IPX1.”</p>

Answer 232:	Please follow Article 6.1.35 of Technical Specifications.
Question 233:	Item 6.1.41, "12 channels ECG measurement cable and electrodes (1 set) for each device" All electrodes can be placed as described in item nos. 6.1.17/18, therefore, we need amendment as following: "2 channels ECG measurement cable for minimum 3 or 5 leads and electrodes (1 set) for each device"
Answer 233:	Please follow Article 6.1.41 of Technical Specifications.
Question 234:	Item 6.1.41, "12 channels ECG measurement cable and electrodes (1 set) for each device" Suggested clause "3 channels ECG measurement cable and electrodes (1 set) for each device"
Answer 234:	Please follow Article 6.1.41 of Technical Specifications.
Question 235:	Item 6.1.44, "Pediatric / child and adult external defibrillation spoons (1 set) for each device" Suggested clause "Pediatric / child and adult external defibrillation paddles (1 set) for each device"
Answer 235:	Please follow Article 6.1.44 of Technical Specifications.
Question 236:	Item 6.1.46, "Adult disposable defibrillation / cardioversion pads (5 pcs) for each device" Suggested clause "Removal from Technical Specification"
Answer 236:	Please follow Article 6.1.46 of Technical Specifications.
Question 237:	Item 6.1.47, "Disposable pediatric / child defibrillation / cardioversion / pads for each device (5 pcs)" Suggested clause "Removal from Technical Specification"
Answer 237:	Please follow Article 6.1.47 of Technical Specifications.
Question 238:	Item 6.1.48, "For each device, the reusable adult SPO2 measurement finger probe and intermediate cable (1 pc)" Suggested clause "Removal from Technical Specification"
Answer 238:	Please follow Article 6.1.48 of Technical Specifications.
Question 239:	Item 6.1.49, "For each device, the reusable pediatric / child SPO2 measurement finger probe and intermediate cable (1 pc)" Suggested clause "Removal from Technical Specification"
Answer 239:	Please follow Article 6.1.49 of Technical Specifications.
Question 240:	Item 6.1.50, "Single-use EtCO2 measurement accessories for each device (5 pcs)"

	Suggested clause “Removal from Technical Specification”
Answer 240:	Please follow Article 6.1.50 of Technical Specifications.
Question 241:	Item 6.1.51, “Reusable adult NIBP cuff and connection hose (1 piece) for each device” Suggested clause “Removal from Technical Specification”
Answer 241:	Please follow Article 6.1.51 of Technical Specifications.
Question 242:	Item 6.1.52, “Reusable pediatric / child NIBP cuff and connection hose (1 piece) for each device” Suggested clause “Removal from Technical Specification”
Answer 242:	Please follow Article 6.1.52 of Technical Specifications.
Question 243:	Item 7.1.16, “The device shall incorporate XDclear probe technology with cool stack technology that improves cool stack penetration and that prevents probe overheating when forced, or the Active array probe technology that contain the circuits associated with the beam former structure on the probe rather than the system, thus offering superior image quality.” Requested amendment “The device shall incorporate XDclear probe or Purewave probe or IQ probe technology thus offering superior image quality.”
Answer 243:	Please follow Article 7.1.16 of Technical Specifications.
Question 244:	Item 7.1.17, “In addition to conventional probe technology, the proposed system shall be able to allow connection of matrix (with 550 crystals at minimum) or x-matrix (with 550 crystals at minimum) or probes with an active array feature.” Requested amendment “In addition to conventional probe technology, the proposed system shall be able to allow connection of Matrix (with 550 crystals at minimum) or X-Matrix (with 550 crystals at minimum) or IQ probe (with minimum 5000 sub-elements per cm ²) features.”
Answer 244:	Please follow Article 7.1.17 of Technical Specifications.
Question 245:	Item 7.1.20, “It shall be possible to connect 3 fully electronic active arrays or pin-free probes simultaneously to the system apart from CW Doppler probe, and the probes to be used shall be selectable using a selector range on the panel. The following probes shall be supplied together with the proposed device.” Requested amendment “It shall be possible to connect 3 fully electronic probes simultaneously to the system and the probes to be used shall be selectable using a selector range on the panel. The following probes shall be supplied together with the proposed device.”
Answer 245:	Please follow Article 7.1.20 of Technical Specifications.

<p>Question 246:</p>	<p>Item 7.1.25, “The system shall feature S-Agile Architecture operating architecture.”</p> <p>Requested amendment “7.1.24 a) The system shall feature S-Agile Architecture or S View-Extended Modular Architecture operating architecture.”</p>
<p>Answer 246:</p>	<p>Please see the Corrigendum No: 2 to Tender Dossier.</p>
<p>Question 247:</p>	<p>Item 7.1.26, “The system shall feature TGC settings as well as Lateral Gain Control (LGC) which allows setting of echo intensity on the lateral line.”</p> <p>Requested amendment “7.1.24 b) The system shall feature TGC settings as well as Lateral Gain Control (LGC) which allows setting of echo intensity on the lateral line.”</p>
<p>Answer 247:</p>	<p>Please see the Corrigendum No: 2 to Tender Dossier.</p>
<p>Question 248:</p>	<p>Item 7.1.27, The system shall be upgradeable with 4D imaging feature if and when desired. When the system is upgraded with such a feature, 4D imaging scan rate shall reach the 30 volume/sec value; furthermore, the system shall also allow installation of dedicated software (autoface reveal etc.) that automatically removes the tissues that prevent imaging the face of fetus.”</p> <p>Requested amendment “7.1.24 c) The system shall be upgradeable with 4D imaging feature if and when desired. When the system is upgraded with such a feature, 4D imaging scan rate shall reach the 30 volume/sec value; furthermore, the system shall also allow installation of dedicated software (Autoface Reveal or Fetal Face or X Light or similar technologies) that automatically removes the tissues that prevent imaging the face of fetus for high quality 3D/4D Imaging .”</p>
<p>Answer 248:</p>	<p>Please see the Corrigendum No: 2 to Tender Dossier.</p>
<p>Question 249:</p>	<p>Item 7.1.28, “The system shall be upgradeable with elastography feature, and the system shall allow use of such feature at 5 probes at minimum; furthermore, the system shall be further upgradeable with shear-wave elastography feature, which is applied with at least one convex and one linear probe.”</p> <p>Requested amendment “7.1.24 d) The system shall be upgradeable with strain elastography feature and the system shall allow use of such feature at 5 probes at minimum or the system shall be upgradeable with shear-wave elastography feature, which is applied with at least one convex and one linear probe.”</p>
<p>Answer 249:</p>	<p>Please see the Corrigendum No: 2 to Tender Dossier.</p>
<p>Question 250:</p>	<p>Item 7.1.29, “The system shall be upgradeable with fusion/navigation feature.”</p> <p>Requested amendment “7.1.24 e) The system shall be upgradeable with fusion/navigation feature.”</p>

Answer 250:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 251:	New additional Item “7.1.24 f) The maximum B Mode frequency of the probes connectable to the system shall be 22.0 MHz or higher.”
Answer 251:	Please follow Article 7.1.24 of Technical Specifications and Please see the Corrigendum No: 2 to Tender Dossier.
Question 252:	Item 7.1.31, “The monitor of the system shall be a high-resolution, vibration-free, 19-inch LCD monitor at minimum.” Requested amendment “7.1.26 The monitor of the system shall be a high-resolution, vibration-free, 19-inch LCD monitor at minimum.The system should also have a touch screen menu display of minimum8.9 inches.”
Answer 252:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 253:	Item 7.1.49, “"Gray scale mapping" in B-mode if the device is over a frozen image; "Priority, color map, color invert, scala baseline" in color Doppler mode; In Pulse Doppler mode, the parameters "invert, sweep speed, angle correction, baseline" must be changed.” Requested amendment “7.1.44 In real time or in frozen mode "Gray scale mapping" in B-mode image, "Color map, color invert, scala baseline" in Color Doppler mode and the parameters "invert, sweep speed, angle correction, baseline" in Pulse Wave Doppler Mode must be capable to be changed by the user.”
Answer 253:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 254:	Item 8.1.6, “Mobile Oxygen Generator, Inspection and Storage System shall be certificated as a single device in a minimum level of II-b, in line with the communique 2012/7 published by the Turkish Medicines and Medical Devices Agency on 14.09.2012; shall bear CE mark and the model of the proposed device shall be registered in the CE certificate and along with the original copy or a notarized copy of the CE certificate, it shall be documented that the liquid and waste gases released by the system do not harm the environment, and to do this, the original or notarized copies of ISO 14001:2015 environment management system certificate shall be submitted along with the letter of proposal.” Requested amendment “Mobile Oxygen Generator, Inspection and Storage System shall be certificated as a single device in a minimum level of II-a, in line with the communique 2012/7 published by the Turkish Medicines and Medical Devices Agency on 14.09.2012; shall bear CE mark and the model of the proposed device shall be registered in the CE certificate and along with the original copy or a notarized copy of the CE certificate, it shall be documented that the liquid and waste gases released by the system do not harm the environment, and to do this, the original or notarized copies of ISO 14001:2015 environment management system certificate shall be submitted along with the letter of proposal.”
Answer 254:	Please see the Corrigendum No: 2 to Tender Dossier.

<p>Question 255:</p>	<p>Item 8.1.9, “System shall consist of;</p> <ul style="list-style-type: none"> • 1 Mobile Oxygen Generator System. • 1 Low pressure compressor • 1 High pressure compressor • 1 Electronic management monitor • 1 Oxygen sensor • Mobile Oxygen Generator System Oxygen tank shall hold 50 Litres. • Mobile Oxygen Generator System shall conform to electrostatic paint and mobile usage.” <p>Requested amendment “System shall consist of;</p> <ul style="list-style-type: none"> • 1 Mobile Oxygen Generator System. • 1 Compressor to produce air in a pressurized way • 1 Air drier which reduces the amount of humidity in the air • 1 Electronic management monitor • 1 Oxygen sensor • Mobile Oxygen Generator System Oxygen tank shall hold 50 Litres. • Mobile Oxygen Generator System shall conform to electrostatic paint and mobile usage.”
<p>Answer 255:</p>	<p>Please see the Corrigendum No: 2 to Tender Dossier.</p>
<p>Question 256:</p>	<p>Item 8.1.10, “Oxygen condensation module shall produce 10 lt/m oxygen at 93 + 3 % density. Oxygen condensation module shall work with ATF system and this system shall be in the form of pressed block consisting of at least 12 columns. Nitrogen (N₂) and other inert gases shall be automatically removed by the module. Module structure shall be non-opening compact. Module shall not be disrupted by variable air and elevation conditions and keep producing oxygen. Oxygen columns shall be of non-oxidising type.”</p> <p>Requested amendment “Oxygen condensation module shall produce 10 lt/m oxygen at 93 + 3 % density.”</p>
<p>Answer 256:</p>	<p>Please see the Corrigendum No: 2 to Tender Dossier.</p>
<p>Question 257:</p>	<p>Item 8.1.11, “Low pressure compressor shall be 100 % oil-free. With a pressure of 2 bar, it shall have the capacity to linearly transmit the air required by the oxygen condensation module.”</p> <p>Requested amendment “Compressor shall be oily or oil-free. It shall have the capacity to linearly transmit the air required by the oxygen condensation module.”</p>
<p>Answer 257:</p>	<p>Please see the Corrigendum No: 2 to Tender Dossier.</p>
<p>Question 258:</p>	<p>Item 8.1.12, “High pressure compressor shall transmit the produced medical oxygen into the oxygen reserve tank or the usage plugs with a pressure of 4-6 bar. High pressure compressor shall have a minimum capacity of 10 lt /m. High pressure compressor shall be 100 % oil-free.”</p> <p>Requested amendment “Compressor shall transmit the produced medical oxygen into the oxygen reserve tank or the usage plugs with a pressure of</p>

	4-6 bar. Oxygen Generator System shall have a minimum capacity of 10 lt /m. Compressor shall be oily or oil-free.”
Answer 258:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 259:	Item 8.1.13, “Electronic Monitor shall have a screen with minimum 2x16 characters. Oxygen purity value, pressure value and operation modes of the system shall be monitored on this screen. System shall run fully automatic. The used system shall be automatic for second use and stay on standby for use.” Requested amendment “Electronic Monitor shall have a screen. Oxygen purity value, pressure value and operation modes of the system shall be monitored on this screen. System shall run fully automatic. The used system shall be automatic for second use and stay on standby for use.”
Answer 259:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 260:	Item 8.1.17, “System shall include a 0,1 micron dust filter 99,999 % micron a bacteria filter subjected to BS EN 13328–1 International personal tests.” Requested amendment “System shall include a 0,1 micron dust filter and a bacteria filter pursuant to Ministry of Health communiqué 2012/7”
Answer 260:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 261:	Item 8.1.19, “System shall be produced from 100 % oil-free materials.” Requested amendment “System must be manufactured from materials suitable for the medical oxygen production system.”
Answer 261:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 262:	Item 9.1.7.5, “The size of each pixel shall be 150 micrometers at most.” Requested amendment “The size of each pixel shall be 175 micrometers at most.”
Answer 262:	Please follow Article 9.1.7.5 of Technical Specifications.
Question 263:	Item 9.1.8.12, “There will be manual and automatic collimator in the system. The collimator will have time-lined field illumination and laser centering.” Requested amendment “There will be manual and automatic collimator in the system. The collimator will have time-lined field illumination and light or laser centering.”
Answer 263:	Please follow Article 9.1.8.12 of Technical Specifications.
Question 264:	Item 9.1.9.2, “The detector block at the chest stand must have a grid, a grid ratio of at least 8:1, a grid frequency of at least 36 lines / cm, and a film focal distance must be adjustable in a wide size and must be automatic collimated according to the selected size, film focal distance should be adjustable between at least 100-180 cm.”

	Requested amendment “The detector block at the chest stand must have a grid, a grid ratio of at least 8:1, a grid frequency of at least 36 lines / cm, and a film focal distance must be adjustable in a wide size and must be collimated according to the selected size, film focal distance should be adjustable between at least 100 cm and 180 cm.”
Answer 264:	Please follow Article 9.1.9.2 of Technical Specifications.
Question 265:	Item 9.1.10.1, “System table tops at least 75 x 220 cm in size and float moving in four directions, transverse a total of at least 26 cm in the longitudinal must move a minimum of 80 cm” Requested amendment “System table tops at least 75 x 220 cm in size and float moving in four directions, transverse a total of at least 23 cm in the longitudinal must move a minimum of 80 cm”
Answer 265:	Please follow Article 9.1.10.1 of Technical Specifications.
Question 266:	Item 9.1.11.1, “The generator must operate with microprocessor control and high frequency or high frequency converter technology. The operating frequency must be at least 25 kHz” Requested amendment “The generator should operate with microprocessor controlled and high frequency technique. The operating frequency must be at least 400 kHz”
Answer 266:	Please follow Article 9.1.11.1 of Technical Specifications.
Question 267:	Item 9.1.11.1, “The generator must operate with microprocessor control and high frequency or high frequency converter technology. The operating frequency must be at least 25 kHz” Requested amendment “The generator must operate with microprocessor control and high frequency or high frequency converter technology. The operating frequency must be at least 100 kHz”
Answer 267:	Please follow Article 9.1.11.1 of Technical Specifications.
Question 268:	Item 9.1.11.3, “The generator power must be at least 64 kW (64 kW = 100 kVp, 640 mA for 0.1 s exposure), current values must be able to be set between at least 10 mA to 640 mA, mAs value must be able to be set between at least 0.1-500 mAs, voltage values must be able to be set between 40 kV and 150 kV.” Requested amendment “The generator power must be at least 50 kW (50 kW = 100 kVp, 500 mA for 0.1 s exposure), current values must be able to be set between at least 10 mA to 630 mA, mAs value must be able to be set between at least 0.1-500 mAs, voltage values must be able to be set between 40 kV and 150 kV.”
Answer 268:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 269:	Item 9.1.11.3, “The generator power must be at least 64 kW (64 kW = 100 kVp, 640 mA for 0.1 s exposure), current values must be able to be set between at least 10 mA to 640 mA, mAs value must be able to be set

	<p>between at least 0.1-500 mAs, voltage values must be able to be set between 40 kV and 150 kV.”</p> <p>Requested amendment “The generator power must be at least 50 kW (50 kW = 100 kVp, 630 mA for 0.1 s exposure), current values must be able to be set between at least 10 mA to 630 mA, mAs value must be able to be set between at least 0.1-500 mAs, voltage values must be able to be set between 40 kV and 150 kV.”</p>
Answer 269:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 270:	<p>Item 9.1.11.3, “The generator power must be at least 64 kW (64 kW = 100 kVp, 640 mA for 0.1 s exposure), current values must be able to be set between at least 10 mA to 640 mA, mAs value must be able to be set between at least 0.1-500 mAs, voltage values must be able to be set between 40 kV and 150 kV.”</p> <p>Requested amendment “The generator power must be at least 64 kW (64 kW = 100 kVp, 640 mA for 0.1 s exposure), current values must be able to be set between at least 25 mA to 640 mA, mAs value must be able to be set between at least 0.5-500 mAs, voltage values must be able to be set between 40 kV and 150 kV”</p>
Answer 270:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 271:	<p>Item 9.1.11.4, “The duration of the radiography should be set to a minimum of 0.001 – 10 seconds.”</p> <p>Requested amendment “The duration of the radiography should be set to a minimum of 0.001 – 5 seconds.”</p>
Answer 271:	Please follow Article 9.1.11.4 of Technical Specifications.
Question 272:	<p>Item 9.1.12.4, “The RAM capacity of the system will be at least 4 GB. The system will be in RAID structure to provide hard disk data security and the capacity will be at least 1 TB.”</p> <p>Requested amendment “The RAM capacity of the system will be at least 4 GB. The system will be in RAID structure to provide hard disk data security and the capacity will be at least 320 GB.”</p>
Answer 272:	Please follow Article 9.1.12.4 of Technical Specifications.
Question 273:	<p>Item 9.1.15.4, “The proposed systems shall not be a assembled system but shall be documented with the proposal that the proposed system shall be presented with UBB records belonging to all brands and models as well as the ISO 13485 certification of the manufacturer's digital x-ray machine and covering all parts of the proposed system on the CE certificate.”</p> <p>Requested amendment “Please remove this item for competition”</p>
Answer 273:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 274:	Item 9.1.15.4, “The proposed systems shall not be a assembled system but shall be documented with the proposal that the proposed system shall be presented with UBB records belonging to all brands and models as well as

	<p>the ISO 13485 certification of the manufacturer's digital x-ray machine and covering all parts of the proposed system on the CE certificate.”</p> <p>Requested amendment “The proposed systems will not be collection systems and will be in charge of a single company. UBB records belonging to alt brands and models of the proposed system must be submitted.”</p>
Answer 274:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 275:	<p>Item 9.1.16.1, “The devices (system) will be guaranteed at least 5 years, and this guarantee will be provided by the manufacturer and authorized representative in Turkey, and the vendor if there is. Maintenance, repair and spare parts shall not be charged at all during the warranty period. The device shall be repaired within 48 hours after a failure notification and within 10 days at the latest, shall be fully operational. Time elapsed while faulty shall not be counted during the warranty period and for every day exceeding the determined period, loss of service due to faults will be applied to the company as a penalty sanction.”</p> <p>Requested amendment “Please change “5 years” to “2 years” for high cost”</p>
Answer 275:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 276:	<p>Item 9.1.16.2, “After the end of the free warranty period, they shall give the annual periodical maintenance fee commitment to the manufacturer and the representative and the authorized dealer, if any, for the spare parts for at least 5 years.”</p> <p>Requested amendment “After the end of the free warranty period, they shall give the annual periodical maintenance fee commitment to the manufacturer and the representative and the authorized dealer, if any, for the spare parts for at least 2 years.”</p>
Answer 276:	Please follow Article 9.1.16.2 of Technical Specifications.
Question 277:	<p>Item 9.1.19 and 9.1.19.1, Lead Coating</p> <p>Requested amendment “Since medical equipment business has different items in turn-key construction work, we are requesting that the room preparation lead coating work be done as a separate purchase. This item has been completely removed”</p>
Answer 277:	Please follow Article 9.1.19 and subtitles of Technical Specifications.
Question 278:	<p>Item 9.1.19 and 9.1.19.1, Lead Coating</p> <p>Requested amendment “Please remove this item for high cost”</p>
Answer 278:	Please follow Article 9.1.19 and subtitles of Technical Specifications.
Question 279:	<p>Item 10.1.4, “Device shall be manufactured microprocessor controlled in line with latest technology; with a 5,5 inch colour touchscreen LCD TFT screen with minimum 800x480 resolution; compact and portable.”</p> <p>We request that the sentence shall be amended as; “Device shall be manufactured microprocessor controlled in line with latest technology; with</p>

	a 4,3 inch colour touchscreen LCD TFT screen with minimum 800x480 resolution; compact and portable.”
Answer 279:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 280:	Item 10.1.4, resolution is required 800 x 480. ECG is not a monitor, so 640 x 480 should be also very comfortable for end user.
Answer 280:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 281:	Item 10.1.4, “Device shall be manufactured microprocessor controlled in line with latest technology; with a 5,5 inch colour touchscreen LCD TFT screen with minimum 800x480 resolution; compact and portable.” Suggested clause “Device shall be manufactured microprocessor controlled in line with latest technology; with a 7 inch colour touchscreen LCD TFT screen with minimum 800x480 resolution; compact and portable.”
Answer 281:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 282:	Item 10.1.4, “Device shall be manufactured microprocessor controlled in line with latest technology; with a 5,5 inch colour touchscreen LCD TFT screen with minimum 800x480 resolution; compact and portable.” Suggested clause “Device shall be manufactured microprocessor controlled in line with latest technology; with a 5,5 inch colour touchscreen LCD TFT screen with minimum 600x480 resolution; compact and portable”
Answer 282:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 283:	Item 10.1.6, “The sampling rate in the device shall be at least 1600 pieces for each channel.” Suggested clause “The sampling rate in the device shall be at least 2000 pieces for each channel.”
Answer 283:	Please follow Article 10.1.6 of Technical Specifications.
Question 284:	Item 10.1.8, Device shall be fed from a 50 Hz grid with 220 V voltage and rechargeable internal battery. We request that the sentence shall be amended as; “Device shall be fed from a 50 Hz grid with 220 V voltage or 220 V rechargeable internal battery.”
Answer 284:	Please follow Article 10.1.8 of Technical Specifications.
Question 285:	Item 10.1.9, “Device shall contain AC network interference filter, DFT filter and EMG filter. It shall be possible for the user to activate and deactivate the filter.” We request that the sentence shall be amended as; “Device shall contain AC network interference filter, DFT filter or EMG filter. The filter settings must be user-configurable.”
Answer 285:	Please see the Corrigendum No: 2 to Tender Dossier.

Question 286:	<p>Item 10.1.9, “Device shall contain AC network interference filter, DFT filter and EMG filter. It shall be possible for the user to activate and deactivate the filter.”</p> <p>Suggested clause “Device shall contain AC network interference filter, DFT filter or EMG filter. It shall be possible for the user to activate and deactivate the filter.”</p>
Answer 286:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 287:	<p>Item 10.1.11, “A connection indicator shall be present to determine electrode disconnection. In addition, the device shall have audio and/or visual warning system for EKG cable disconnection, lowered charge in battery, out of paper in logger, pulse rate exceeding or dropping below the set alarm limit etc.”</p> <p>Suggested clause “A connection indicator shall be provided to disconnect the electrode connection. In addition, the device shall provide audible or visual warning to the audio and / or visual warning system for the ECG cable break, battery charge shortage, paper run out condition.”</p>
Answer 287:	Please follow Article 10.1.11 of Technical Specifications.
Question 288:	<p>Item 10.1.12, “Amplitude of device shall be possible to be set to 5 mm/mV, 10 mm/mV, 20 mm/mV. Calibration voltage must be adjusted to 10 mm 1 mV.”</p> <p>Suggested clause “Amplitude of device shall be possible to be set to 5 mm/mV, 10 mm/mV, 20 mm/mV.”</p>
Answer 288:	Please follow Article 10.1.12 of Technical Specifications.
Question 289:	<p>Item 10.1.15, “Device shall have an alphanumeric keyboard as aa standard. Alphanumeric keyboard shall consist of membrane keys.”</p> <p>Suggested clause “Device shall have an alphanumeric or touch sensitive keyboard as a standard.”</p>
Answer 289:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 290:	<p>Item 10.1.15, “Device shall have an alphanumeric keyboard as aa standard. Alphanumeric keyboard shall consist of membrane keys.”</p> <p>Suggested clause “The device shall be able to output 12 channels at the same time. It shall have an alphanumeric keyboard.”</p>
Answer 290:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 291:	<p>Item 10.1.16, “The alphanumeric keypad on the device shall contain on-off, start-stop, mm/s set button, mm/mv set button, lead selection, filter set button, auto-manual pull mode selection button, exit, enter keys.”</p> <p>Suggested clause “The device shall be capable of opening and closing, starting-stopping, mms setting, mm / mV setting, filet setting, automatic manual pulling and exit workshops.”</p>

Answer 291:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 292:	Item 10.1.16, “The alphanumeric keypad on the device shall contain on-off, start-stop, mm/s set button, mm/mv set button, lead selection, filter set button, auto-manual pull mode selection button, exit, enter keys.” Suggested clause “The alphanumeric or touch sensitive keypad on the device shall contain on-off, start-stop, mm / s setting button, mm / mv setting button, lead selection, filter setting button, auto-manual pull mode selection button, exit, enterkeys.”
Answer 292:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 293:	Item 10.1.17, “Device shall log on thermal paper with a minimum dimension of 110 mm. paper can be roll-type or Z-folded type.” Suggested clause “Device shall log on thermal paper with a minimum dimension of 210 mm. paper can be roll-type or Z-folded type. It shall be able to use the normal facsimile paper.”
Answer 293:	Please follow Article 10.1.17 of Technical Specifications.
Question 294:	Item 10.1.18, “Device shall store at least 150 EKG logs in its standard internal memory. Data stored in memory should not be lost in case of power cut or if battery is out. Data stored in memory should be possible to call back on the screen, and logged on paper. As an option, it shall be possible for EKG data and medical data to be sent directly from the device to an e-mail address.” Suggested clause “Device shall store at least 150 EKG logs in its standard internal memory. Data stored in memory shall not be lost in case of power cut or if battery is out. Data stored in memory shall be possible to call back on the screen, and logged on paper. As an option, EKG data and medical data shall be able to be sent directly from the device to an e-mail address.
Answer 294:	Please follow article 10.1.18 of Technical Specifications.
Question 295:	Item 10.1.18, “Device shall store at least 150 EKG logs in its standard internal memory. Data stored in memory should not be lost in case of power cut or if battery is out. Data stored in memory should be possible to call back on the screen, and logged on paper. As an option, it shall be possible for EKG data and medical data to be sent directly from the device to an e-mail address.” We kindly request the sentence at the below to be added to the mentioned Article; Because of the Migrant Health Centers will use the Ministry of Health software program with in themselves and other Migrant Health Centers, the device must send the results of the shots to the HIS or PC address to prevent paper usage and allow the doctor to monitor the chart clearly.”
Answer 295:	Please follow article 10.1.18 of Technical Specifications.

Question 296:	Item 10.1.20, “At least 1 USB host, 1 USB device, 1 RJ-45 Ethernet socket should be standard in the device.” Suggested clause “As a standard, the device shall have at least 1 USB host or 1 USB device or 1 RJ-45 Ethernet socket.”
Answer 296:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 297:	Item 10.1.21, please accept “at least 30 logs”. 70 log is so high and specific.
Answer 297:	Please follow Article 10.1.21 of Technical Specifications.
Question 298:	Item 10.1.22, please accept the maximum weight up to 3 Kg, in order to increase competition. 2 Kg is so low and specific.
Answer 298:	Please follow Article 10.1.22 of Technical Specifications.
Question 299:	Item 10.1.22, “Maximum weight of device shall not exceed 2 kg.” Suggested clause “Maximum weight of device shall not exceed 4 kg.”
Answer 299:	Please follow Article 10.1.22 of Technical Specifications.
Question 300:	Item 10.1.23, “Maximum dimensions of device shall be 350 mm x 250 mm x 50 mm (L x W x H).” Please accept the tolerance as “350 mm ±10% x 250 mm ±10% x 50 – 75 mm”.
Answer 300:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 301:	Item 10.1.23, “Maximum dimensions of device shall be 350 mm x 250 mm x 50 mm (L x W x H).” We request that the sentence shall be amended as; “Maximum dimensions of device shall be 396 mm x 290 mm x 65 mm.” (The mentioning of the device measure and weight prevent the equal competition.)”
Answer 301:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 302:	Item 10.1.23, “Maximum dimensions of device shall be 350 mm x 250 mm x 50 mm (L x W x H).” Suggested clause, “Maximum dimensions of deviceshall be 350 mm x 250 mm x 85 mm (L x W x H)”
Answer 302:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 303:	Item 10.1.24, “As an option, device shall be connected with HIS in HL7 (Health Level Seven) work order recognition system standard.” Suggested clause “As an option, device shall be connected with HIS in HL7 (Health Level Seven) or DICOM work order recognition system standard.”
Answer 303:	Please follow article 10.1.24 of Technical Specifications.
Question 304:	Item 10.1.25, “Medical data should be recorded with at least two of SCP, PDF, XML formats.”

	Suggested clause “Medical data shall be possible to be logged with at least one of SCP, PDF, XML formats.”
Answer 304:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 305:	Item 10.1.26, “Pressure density shall be adjusted in 3 levels.” Suggested clause “We suggest that this clause is removed.”
Answer 305:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 306:	Item 10.1.26, “Pressure density shall be adjusted in 3 levels.” We kindly request this Article should be removed because there is no benefit to the user or the institution in usage and result output.
Answer 306:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 307:	Item 10.1.26, “Pressure density shall be adjusted in 3 levels.” It will be appropriate to remove this article from the technical specification since the pressure intensity is indicated in 3 levels causing competition to be decreases because there are devices in different models in the market
Answer 307:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 308:	Item 10.1.27, “Spirometric examination should be possible.” Suggested clause “Removal from technical specification”
Answer 308:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 309:	Item 10.1.27, do you want to make spirometric measurement? If yes, there should be added both software and hardware. Please clarify that will you make spirometric measurement or not.”
Answer 309:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 310:	Item 11.1.7, “The maximum speed of the device should be at least 5000 rpm and the maximum RCF value should be at least 2000 g.” What is important in centrifugation process is the RCF value exerted on the sample. As the RCF value is calculatated as the product of the rotor radius and the RPM value, manufacturers can reach the requested RCF values with different combinations of radius and RPM. Therefore, in order to favour competition, we request that the sentence shall be amended as: “The maximum speed of the device should be at least 4100 rpm and the maximum RCF value should be at least 2000 g.”
Answer 310:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 311:	Item 11.1.13, “Manufacturers EN ISO 9001: 2008 in accordance with ISO 13485: 2013 Quality Management System.” The certification process is performed in accordance with ISO 13485:2012 until 2019. In order to favour competition, we request that this standard is amended as

	“Manufacturers EN ISO 9001: 2008 in accordance with ISO 13485: 2012 Quality Management System.”
Answer 311:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 312:	Item 12.1.8, “Digital X-Ray CMOS Sensor should have maximum 19µm Pixel size (1 (One) Sensor)” Suggested clause “Digital X-Ray CMOS Sensor should have maximum 20µm Pixel size (1 (One) Sensor)”
Answer 312:	Please follow Article 12.1.8 of Technical Specifications.
Question 313:	Item 12.1.9, “Digital X-Ray Sensor 37 x 25 mm Should not be bigger than outer dimensions (1 (One) Sensor)” Suggested clause “Digital X-Ray Sensor 39 x 25 mm Should not be bigger than outer dimensions (1 (One) Sensor)”
Answer 313:	Please follow Article 12.1.9 of Technical Specifications.
Question 314:	Item 12.1.10, “Digital X-Ray Sensor (30.02 x 19,95) 600 mm ² The active area should have 1.659 Mega Pixels (1050 x 1580) (1 (One) Sensor)” Suggested clause “Digital X-Ray Sensor (30 x 20) 600 mm ² The active area should have 1500 Mega Pixels (1500 x 1000) (1 (One) Sensor)”
Answer 314:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 315:	Item 12.2.4.10, “Unit combination shall be painted with preventive electrostatic oven-drying paint against corrosion.” Suggested clause “Unit combination shall be painted with preventive electrostatic or acrylic oven-drying paint against corrosion.”
Answer 315:	Please follow Article 12.2.4.10 of Technical Specifications.
Question 316:	Item 12.2.9.10, “Chair programs; shall be possible to be commanded from 3 different points, namely tablet main panel, foot pedal and assistant command panel.” Suggested clause “Chair programs; shall be possible to be commanded from 2 different points, namely tablet main panel and foot pedal or assistant command panel.”
Answer 316:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 317:	Item 12.2.7.1, “Spittoon bowl shall be produced from ceramic, porcelain, enamel or opal glass and easy to remove and wipe.” Suggested clause “Spittoon bowl shall be produced from ceramic, porcelain, enamel or opal glass and easy to wipe.”
Answer 317:	Please follow Article 12.2.7.1 of Technical Specifications.
Question 318:	Item 12.2.7.5, “The cup-holder and spittoon washing pipes on the spittoon shall be removable for cleaning purposes.”

	Suggested clause “The cup-holder and spittoon washing on the spittoon shall be easy accessible for cleaning purposes.”
Answer 318:	Please follow Article 12.2.7.5 of Technical Specifications.
Question 319:	New additional Item “The spittoon should be comfortable to use and hygienic to the patient and should be in one-touch spitting position with one key.”
Answer 319:	Please follow Article 12.2.7 and subtitles of Technical Specifications.
Question 320:	Item 12.9.8, “The autoclave should have a volume of at least 22 l. The metal load capacity is at least 5 kg, the textile load capacity is at least 1.8 kg.” We request to revise the metal load capacity as minimum 4 kg textile capacity as 1.2 kg.
Answer 320:	Please follow Article 12.9.8 of Technical Specifications.
Question 321:	Item 12.9.9, “The dimensions of the device should be minimum 445 x 410 x 620 cm. The maximum weight shall not exceed 56 kg. Sound level should not exceed 53 dB.” We request to revise sound level as 66 DB.
Answer 321:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 322:	Item 12.9.9, “The dimensions of the device should be minimum 445 x 410 x 620 cm. The maximum weight shall not exceed 56 kg. Sound level should not exceed 53 dB.” According to EN 13060 standard, the maximum volume of benchtop sterilizer cannot exceed 60 liters. As the given dimensions are in cm, the volume of the resulting device is beyond the acceptable limits. Generally, the benchtop steam sterilizers are preferred to have a cylindrical chamber with a volume between 20-22 liters. Additionally, EN 13060 standard defines the maximum sound level as 60 dB. Therefore, we request you to amend this sentence as: “The inner chamber volume should be minimum 22 liters. The maximum weight shall not exceed 56 kg. Sound level should not exceed 60 dB.”
Answer 322:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 323:	Item 12.9.11, “It shall be an original cage in the bowl and the original should be suitable for use with at least 5 trays or 3 trays. There shall be at least 5 aluminum trays out of the device.” We want to revise number of trays as minimum 4 trays.
Answer 323:	Please follow Article 12.9.11 of Technical Specifications.
Question 324:	Item 12.9.12, “It shall be able to sterilize all solid, hollow, hollow B and textile products defined in EN 13060 standards, with or without pouches. This should be done in all programs.”

	<p>EN 13060 standard defines the materials to be sterilized, their sterilization temperatures and corresponding durations. All materials can not be sterilized in all programs because different materials have different temperature resistance.</p> <p>Therefore, we request you to amend this sentence as:</p> <p>“It shall be able to sterilize all solid, hollow, hollow B and textile products as defined in EN 13060 standards”</p>
Answer 324:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 325:	<p>Item 12.9.13, “The device shall have a touch-sensitive display on it, the device's control shall be touchable from this screen.”</p> <p>We request to change Touch screen as Touch screen or button screen.</p>
Answer 325:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 326:	<p>Item 12.9.13, “The device shall have a touch-sensitive display on it, the device's control shall be touchable from this screen.”</p> <p>All manufactures do not have touch screen displays but also TFT LCD displays.</p> <p>Therefore, in order to favour competition, we request that the sentence shall be amended as:</p> <p>“The device shall have a touch-sensitive or LCD display on it, the device's control shall be controlled from this screen”</p>
Answer 326:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 327:	<p>Item 12.9.14, “At least 4 liters of clean and at least 4 liters of dirty water tank capacity, clean water tank should be available for at least 8-12 programs. Warning message should be displayed on the screen of the device when the empty water tank is empty or the waste water tank is full.”</p> <p>We request to revise clean and dirty water tanks capacity as 3.5 lt.</p>
Answer 327:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 328:	<p>Item 12.9.14, “At least 4 liters of clean and at least 4 liters of dirty water tank capacity, clean water tank should be available for at least 8-12 programs. Warning message should be displayed on the screen of the device when the empty water tank is empty or the waste water tank is full.”</p> <p>Different manufacturers’ products have different tank capacities. In addition, the number of sterilization cycles that can be performed by one clean tank capacity depends highly on the program selected.</p> <p>Therefore, in order to favour competition, we request that the sentence shall be amended as:</p> <p>“At least 4 liters of clean and at least 3 liters of dirty water tank capacity, clean water tank should be available for at least 5 programs.</p> <p>Warning message should be displayed on the screen of the device when the empty water tank is empty or the waste water tank is full.”</p>

Answer 328:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 329:	<p>Item 12.9.16, “The device shall also have Bowie & Dick, Helix and Vacuum Test programs and these test programs should be selectable from the touch screen.”</p> <p>As indicated in Item 12.9.13, this item should be amended as “The device shall also have Bowie & Dick, Helix and Vacuum Test programs and these test programs should be selectable”</p>
Answer 329:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 330:	<p>Item 12.9.17, “The device's power button, tank drain connections and bacteria filter shall be located on the device and hidden with a cap.”</p> <p>There is not a such requirement in EN 13060. Depending on the design, the manufacturers may locate the power button, tank drain connections and bacteria filter on easily reachable place.</p> <p>Therefore, in order to favour competition, we request that the sentence shall be amended as: “The device's power button, tank drain connections and bacteria filter shall be located on an easily reachable location on the device”</p>
Answer 330:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 331:	<p>Item 12.9.18, There shall be a meter counting cycles for bacterial filtration, lid seal, and general service, and the value of the three counters at the end of each cycle shall also be reduced. One of the counters should give a warning message on the screen when zero is reached. The bacterial filter should be at least 400 and the lid seal should be replaced after at least 1000 applications.</p> <p>Different manufacturers’ products have different periods for the replacement of filter and the gasket. To increase competition, this item should be amended as:</p> <p>There shall be a meter counting cycles for bacterial filtration, lid seal, and general service, and the value of the three counters at the end of each cycle shall also be reduced. One of the counters should give a warning message on the screen when zero is reached. The bacterial filter should be at least 300 and the lid seal should be replaced after at least 500 applications.</p>
Answer 331:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 332:	<p>Item 12.9.18, There shall be a meter counting cycles for bacterial filtration, lid seal, and general service, and the value of the three counters at the end of each cycle shall also be reduced. One of the counters should give a warning message on the screen when zero is reached. The bacterial filter should be at least 400 and the lid seal should be replaced after at least 1000 applications.</p> <p>We want to revise “lid seal should be replaced after at least 1000 applications” as lid seal should be replaced after at least 990 applications”</p>

Answer 332:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 333:	Item 13.11.4, “Shall be compatible with the cables in use.” The cables that are in use or will be used needs to be clarified.
Answer 333:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 334:	Item 14.7, dimensions of the sture is not specifed. A specification needs to be mad efor this item.
Answer 334:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 335:	Item 14.10.8, “Shall be 5.5 CM x 10 M long.” The global dimension for cloth plaster is 5 cm x 10 cm. the change of 5.5 cm to 5 cm is required.
Answer 335:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 336:	Item 14.26.5, “The dimensions shall be 10 x 10 (± 0.5 cm).” The dimension of the product shall be 8.5 x 11.5 cm. The 10 x 10 dimension is related to just one firm who is violating the tender term of origin.
Answer 336:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 337:	Item 14.34 and 14.35, There are missing information on which kind of bandages they are.
Answer 337:	Please see the Corrigendum No: 2 to Tender Dossier.
Question 338:	Item 14.42, There are missing information that needs to specify which type of mask it is.
Answer 338:	Please see the Corrigendum No: 2 to Tender Dossier.