

**CORRIGENDUM No: 5**  
**to the TENDER DOSSIER**

**Publication Ref: SIHHAT/2018/SUP/INT/10**

**Subject: Supply of Medical Equipment for the Secondary Healthcare Premises**

**Location –Europe (non EU/Turkey)**

The Tender Dossier is corrected/modified as follows:

**ANNEX II+III: TECHNICAL SPECIFICATIONS + TECHNICAL OFFER**

**Instead of (the former text):**

Item Number **10.2.9** (Corrigendum No: 3)

...

i) Endotracheal and tracheostomy "Tube Resistance Compensation" (TRC/ATC/ARC/TC) shall be available in the device. So that, the user can directly adjust the Tube Resistance Compensation and diameter, compensation percentage or the device shall have the feature of circuit compliance compensation that can be turned on and off.

**Read (new text):**

Item Number **10.2.9**

...

i) Endotracheal and tracheostomy "Tube Resistance Compensation" (TRC/ATC/ARC/TC) shall be available in the device. So that, the user can directly adjust the Tube Resistance Compensation and diameter, compensation percentage or the device shall have the feature of circuit compliance compensation.